

Data Path : Z:\HPCHEM1\BNA E\DATA\BE030617\
 Data File : BE092786.D
 Acq On : 6 Mar 2017 14:00
 Operator : SJ/MA
 Sample : SSTDICC005
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC005

Manual Integrations
 APPROVED

mohammad
 3/7/2017 4:22:49 PM

Quant Time: Mar 06 15:32:31 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE030617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 06 14:29:53 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.09	152	11196	5.00	ng	0.00
7) Naphthalene-d8	10.88	136	51071	5.00	ng	0.00
12) Acenaphthene-d10	14.75	164	34193	5.00	ng	0.00
18) Phenanthrene-d10	17.51	188	68892	5.00	ng	0.00
24) Chrysene-d12	21.68	240	91497	5.00	ng	0.00
31) Perylene-d12	24.01	264	81584	5.00	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.60	112	12653	5.81	ng	-0.01
5) Phenol-d6	7.27	99	17565	6.57	ng	-0.02
8) Nitrobenzene-d5	9.27	82	17546	6.17	ng	-0.02
13) 2,4,6-Tribromophenol	16.25	330	5607	7.73	ng	-0.01
14) 2-Fluorobiphenyl	13.36	172	41190	4.94	ng	-0.02
26) Terphenyl-d14	20.12	244	58993	5.05	ng	0.00

Target Compounds				Qvalue		
2) 1,4-Dioxane	3.33	88	6589	4.14	ng	94
3) n-Nitrosodimethylamine	3.68	42	9971	6.02	ng	98
6) bis(2-Chloroethyl)ether	7.52	93	15358	5.19	ng	# 27
9) Naphthalene	10.91	128	55233	5.00	ng	# 95
10) Hexachlorobutadiene	11.20	225	7899	4.86	ng	99
11) 2-Methylnaphthalene	12.54	142	37617	5.42	ng	97
15) Acenaphthylene	14.46	152	273671	5.30	ng	100
16) Acenaphthene	14.80	154	39026	5.20	ng	92
17) Fluorene	15.80	166	51989	5.40	ng	98
19) Hexachlorobenzene	16.82	284	15210	5.72	ng	# 66
21) Phenanthrene	17.53	178	86663	5.29	ng	# 93
22) Anthracene	17.63	178	89320	5.93	ng	98
23) Fluoranthene	19.54	202	421709	6.01	ng	98
25) Pyrene	19.90	202	452516	5.19	ng	100
27) Benzo(a)anthracene	21.66	228	488451m	5.50	ng	
28) Chrysene	21.70	228	400579m	3.81	ng	
29) Bis(2-ethylhexyl)phthalate	21.59	149	47912	3.89	ng	# 67
30) Indeno(1,2,3-cd)pyrene	26.44	276	469999	3.96	ng	97
32) Benzo(b)fluoranthene	23.29	252	115491	5.01	ng	# 97
33) Benzo(k)fluoranthene	23.34	252	108462	4.45	ng	94
34) Benzo(a)pyrene	23.90	252	106448	5.08	ng	# 95
35) Dibenzo(a,h)anthracene	26.48	278	99027	4.48	ng	94
36) Benzo(g,h,i)perylene	27.21	276	408704	4.37	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA E\DATA\BE030617\
Data File : BE092786.D
Acq On : 6 Mar 2017 14:00
Operator : SJ/MA
Sample : SSTDICC005
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:49 PM

Quant Time: Mar 06 15:32:31 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE030617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 06 14:29:53 2017
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 8.09 min Scan# 573

Delta R.T. -0.00 min

Lab File: BE092786.D

Acq: 6 Mar 2017 14:00

Instrument :

BNA_E

ClientSampled :

SSTDIC005

Tot Ion:152 Resp: 11196

Ion Ratio Lower Upper

152 100


150 151.0 106.0 159.0

115 52.4 31.2 46.8#

Manual Integrations
APPROVED

mohammad




3/7/2017 4:22:49 PM



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 4.14 ng

RT: 3.33 min Scan# 64

Delta R.T. 0.00 min

Lab File: BE092786.D

Acq: 6 Mar 2017 14:00


Tgt Ion: 88 Resp: 6589

Ion Ratio Lower Upper

88 100

58 86.6 63.6 95.4



43 36.3 28.0 42.0






Abundance Ion 88.00 (87.70 to 88.70): BE0

Ion 58.00 (57.70 to 58.70): BE0

Ion 43.00 (42.70 to 43.70): BE0





#3

n-Nitrosodimethylamine

Concen: 6.02 ng

RT: 3.68 min Scan# 94

Delta R.T. -0.02 min

Lab File: BE092786.D

Acq: 6 Mar 2017 14:00

Tot Ion:	42	Resp:	9971
Ion	Ratio	Lower	Upper
42	100		
74	105.4	86.0	129.0
44	6.9	5.1	7.7

Instrument :





BNA_E

ClientSampleId :

SSTDIC005

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:49 PM



#4

2-Fluorophenol

Concen: 5.81 ng


RT: 5.60 min Scan# 350


Delta R.T. -0.01 min

Lab File: BE092786.D

Acq: 6 Mar 2017 14:00

Tgt Ion:	112	Resp:	12653
Ion	Ratio	Lower	Upper
112	100		
64	68.3	53.5	80.3
63	36.2	27.9	41.9





#5

Phenol-d6

Concen: 6.57 ng

RT: 7.27 min Scan# 537

Delta R.T. -0.02 min

Lab File: BE092786.D

Acq: 6 Mar 2017 14:00

Instrument :

BNA_E





ClientSampleId :

SSTDIC005

Tot Ion	Ratio	Lower	Upper
99	100		
42	0.0	16.0	24.0#
71	31.9	30.6	45.8

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:49 PM



#6

bis(2-Chloroethyl)ether

Concen: 5.19 ng




RT: 7.52 min Scan# 548


Delta R.T. -0.00 min

Lab File: BE092786.D

Acq: 6 Mar 2017 14:00

Tgt Ion	Ratio	Lower	Upper
93	100		
63	0.0	71.6	107.4#
95	34.2	24.0	36.0





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.88 min Scan# 704

Delta R.T. 0.00 min

Lab File: BE092786.D

Acq: 6 Mar 2017 14:00

Instrument :

BNA_E


ClientSampleId :

SSTDIC005

Tot Ion:	136	Resp:	51071
Ion Ratio	Lower	Upper	
136	100		
137	11.6	8.2	12.4
54	5.7	9.6	14.4#
68	4.4	6.7	10.1#

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:49 PM




Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E


Ion 54.00 (53.70 to 54.70): BE0

Ion 68.00 (67.70 to 68.70): BE0

Time-->



Time-->



#8

Nitrobenzene-d5

Concen: 6.17 ng


RT: 9.27 min Scan# 624

Delta R.T. -0.02 min

Lab File: BE092786.D

Acq: 6 Mar 2017 14:00

Tgt Ion:	82	Resp:	17546
Ion Ratio	Lower	Upper	
82	100		
128	62.9	39.0	58.4#
54	36.1	44.8	67.2#




Abundance Ion 82.00 (81.70 to 82.70): BE0


Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0

Time-->



Time-->



#9

Naphthalene

Concen: 5.00 ng

RT: 10.91 min Scan# 706

Delta R.T. -0.02 min

Lab File: BE092786.D

Acq: 6 Mar 2017 14:00

Instrument :

BNA_E





ClientSampleId :

SSTDIC005

Tot Ion	Ratio	Lower	Upper
128	100		
129	10.6	11.2	16.8#
127	13.6	11.6	17.4

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:49 PM



#10

Hexachlorobutadiene

Concen: 4.86 ng




RT: 11.20 min Scan# 721


Delta R.T. 0.00 min

Lab File: BE092786.D

Acq: 6 Mar 2017 14:00

Tgt Ion	Ratio	Lower	Upper
225	100		
223	63.8	50.6	76.0
227	63.3	51.4	77.0





#11

2-Methylnaphthalene

Concen: 5.42 ng

RT: 12.54 min Scan# 807

Delta R.T. -0.03 min

Lab File: BE092786.D

Acq: 6 Mar 2017 14:00

Instrument :

BNA_E

ClientSampleId :





SSTDIC005

Tot Ion:142 Resp: 37617

Ion	Ratio	Lower	Upper
142	100		
141	94.2	73.7	110.5
115	40.1	34.0	51.0

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:49 PM



#12

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.75 min Scan# 976




Delta R.T. 0.00 min


Lab File: BE092786.D

Acq: 6 Mar 2017 14:00

Tgt Ion:164 Resp: 34193

Ion	Ratio	Lower	Upper
164	100		
162	104.0	71.4	107.0
160	46.0	27.4	41.2






#13
 2,4,6-Tribromophenol
 Concen: 7.73 ng
 RT: 16.25 min Scan# 1083
 Delta R.T. -0.01 min
 Lab File: BE092786.D
 Acq: 6 Mar 2017 14:00

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC005

Tot Ion	Ratio	Lower	Upper
330	100		
332	96.8	75.4	113.0
141	43.7	31.0	46.6

Manual Integrations
 APPROVED

mohammad
 3/7/2017 4:22:49 PM






Abundance

Ion 329.80 (329.50 to 330.50): E


Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E



#14
 2-Fluorobiphenyl
 Concen: 4.94 ng
 RT: 13.36 min Scan# 878
 Delta R.T. -0.02 min
 Lab File: BE092786.D
 Acq: 6 Mar 2017 14:00

Tgt Ion	Ratio	Lower	Upper
172	100		
171	33.7	30.1	45.1
170	22.1	21.8	32.6






Abundance

Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E





#15

Acenaphthylene

Concen: 5.30 ng

RT: 14.46 min Scan# 959

Delta R.T. 0.00 min

Lab File: BE092786.D

Acq: 6 Mar 2017 14:00

Instrument :

BNA_E





ClientSampleId :

SSTDIC005

Tot Ion:	152	Resp:	273671
Ion Ratio	Lower	Upper	
152	100		
151	4.8	3.9	5.9
153	13.1	10.4	15.6

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:49 PM



#16

Acenaphthene

Concen: 5.20 ng




RT: 14.80 min Scan# 979


Delta R.T. -0.02 min

Lab File: BE092786.D

Acq: 6 Mar 2017 14:00

Tgt Ion:	154	Resp:	39026
Ion Ratio	Lower	Upper	
154	100		
153	457.2	350.8	526.2
152	229.1	171.1	256.7





#17

Fluorene

Concen: 5.40 ng

RT: 15.80 min Scan# 1044

Delta R.T. -0.01 min

Lab File: BE092786.D

Acq: 6 Mar 2017 14:00

Instrument :

BNA_E





ClientSampleId :

SSTDIC005

Tot Ion:	166	Resp:	51989
Ion Ratio	Lower	Upper	
166	100		
165	99.6	78.2	117.4
167	13.5	11.0	16.4

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:49 PM



#18

Phenanthrene-d10

Concen: 5.00 ng




RT: 17.51 min Scan# 1151


Delta R.T. -0.00 min

Lab File: BE092786.D

Acq: 6 Mar 2017 14:00

Tgt Ion:	188	Resp:	68892
Ion Ratio	Lower	Upper	
188	100		
94	2.9	3.2	4.8#
80	2.3	2.6	3.8#









#19
Hexachlorobenzene
Concen: 5.72 ng
RT: 16.82 min Scan# 1121
Delta R.T. -0.00 min
Lab File: BE092786.D
Acq: 6 Mar 2017 14:00

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion	Ratio	Lower	Upper
284	100		
142	0.0	29.1	43.7#
249	31.9	27.9	41.9




Manual Integrations
APPROVED


mohammad
3/7/2017 4:22:49 PM



#21
Phenanthrene
Concen: 5.29 ng
RT: 17.53 min Scan# 1152
Delta R.T. -0.02 min
Lab File: BE092786.D
Acq: 6 Mar 2017 14:00

Tgt Ion	Ratio	Lower	Upper
178	100		
176	20.6	15.8	23.8
179	14.7	16.0	24.0#





#22

Anthracene

Concen: 5.93 ng

RT: 17.63 min Scan# 1156

Delta R.T. -0.02 min

Lab File: BE092786.D

Acq: 6 Mar 2017 14:00

Instrument :

BNA_E

ClientSampleId :

SSTDIC005

Tot Ion:178 Resp: 89320

Ion Ratio Lower Upper

178 100





176 19.8 15.0 22.4

179 14.6 12.2 18.2

Manual Integrations
APPROVED

mohammad

3/7/2017 4:22:49 PM



#23

Fluoranthene

Concen: 6.01 ng

RT: 19.54 min Scan# 1262

Delta R.T. -0.02 min

Lab File: BE092786.D

Acq: 6 Mar 2017 14:00




Tgt Ion:202 Resp: 421709


Ion Ratio Lower Upper

202 100

101 2.9 2.2 3.4

203 18.0 13.7 20.5





#24

Chrysene-d12

Concen: 5.00 ng

RT: 21.68 min Scan# 1372

Delta R.T. 0.00 min

Lab File: BE092786.D

Acq: 6 Mar 2017 14:00

Instrument :

BNA_E

ClientSampleId :

SSTDIC005

Tot Ion:240 Resp: 91497

Ion Ratio Lower Upper

240 100


120 4.2 2.6 4.0#

236 30.1 24.6 37.0

Manual Integrations
APPROVED

mohammad




3/7/2017 4:22:49 PM



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#25

Pyrene

Concen: 5.19 ng

RT: 19.90 min Scan# 1283

Delta R.T. -0.02 min

Lab File: BE092786.D

Acq: 6 Mar 2017 14:00


Tgt Ion:202 Resp: 452516

Ion Ratio Lower Upper

202 100

200 5.4 4.2 6.4



203 17.6 13.9 20.9




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E









#26
 Terphenyl-d14
 Concen: 5.05 ng
 RT: 20.12 min Scan# 1296
 Delta R.T. -0.00 min
 Lab File: BE092786.D
 Acq: 6 Mar 2017 14:00

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC005

Tot Ion	Ratio	Lower	Upper
244	100		
212	6.9	6.7	10.1
122	1.7	3.6	5.4#




Manual Integrations
 APPROVED


mohammad
 3/7/2017 4:22:49 PM



#27
 Benzo(a)anthracene
 Concen: 5.50 ng m
 RT: 21.66 min Scan# 1371
 Delta R.T. 0.00 min
 Lab File: BE092786.D
 Acq: 6 Mar 2017 14:00

Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.3	23.6	35.4
229	14.7	13.3	19.9









#28
Chrysene
Concen: 3.81 ng/ml
RT: 21.70 min Scan# 1373
Delta R.T. -0.02 min
Lab File: BE092786.D
Acq: 6 Mar 2017 14:00

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion	Ratio	Lower	Upper
228	100		
226	49.2	36.3	54.5
229	10.4	10.6	16.0




Manual Integrations
APPROVED


mohammad
3/7/2017 4:22:49 PM



#29
Bis(2-ethylhexyl)phthalate
Concen: 3.89 ng/ml
RT: 21.59 min Scan# 1368
Delta R.T. 0.00 min
Lab File: BE092786.D
Acq: 6 Mar 2017 14:00

Tgt Ion	Ratio	Lower	Upper
149	100		
167	32.4	16.0	24.0
279	2.3	16.0	24.0









#30
Indeno(1,2,3-cd)pyrene
Concen: 3.96 ng
RT: 26.44 min Scan# 1707
Delta R.T. -0.03 min
Lab File: BE092786.D
Acq: 6 Mar 2017 14:00

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion	Ratio	Lower	Upper
276	100		
138	28.5	20.3	30.5
277	24.8	19.7	29.5




Manual Integrations
APPROVED


mohammad
3/7/2017 4:22:49 PM



#31
Perylene-d12
Concen: 5.00 ng
RT: 24.01 min Scan# 1537
Delta R.T. 0.00 min
Lab File: BE092786.D
Acq: 6 Mar 2017 14:00

Tgt Ion	Ratio	Lower	Upper
264	100		
260	26.3	20.1	30.1
265	20.9	17.9	26.9









#32
Benzo(b)fluoranthene
Concen: 5.01 ng
RT: 23.29 min Scan# 1475
Delta R.T. 0.00 min
Lab File: BE092786.D
Acq: 6 Mar 2017 14:00

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion	Ratio	Lower	Upper
252	100		
253	23.1	18.7	28.1
125	10.0	10.5	15.7




Manual Integrations
APPROVED


mohammad
3/7/2017 4:22:49 PM



#33
Benzo(k)fluoranthene
Concen: 4.45 ng
RT: 23.34 min Scan# 1479
Delta R.T. 0.00 min
Lab File: BE092786.D
Acq: 6 Mar 2017 14:00

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.3	20.3	30.5
125	10.1	9.6	14.4









#34
Benzo(a)pyrene
Concen: 5.08 ng
RT: 23.90 min Scan# 1528
Delta R.T. 0.00 min
Lab File: BE092786.D
Acq: 6 Mar 2017 14:00

Instrument :
BNA_E
ClientSampleId :
SSTDICC005

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.2	19.0	28.6
125	11.4	11.8	17.8




Manual Integrations
APPROVED


mohammad
3/7/2017 4:22:49 PM



#35
Dibenzo(a,h)anthracene
Concen: 4.48 ng
RT: 26.48 min Scan# 1709
Delta R.T. -0.02 min
Lab File: BE092786.D
Acq: 6 Mar 2017 14:00

Tgt Ion	Ratio	Lower	Upper
278	100		
139	14.0	13.8	20.8
279	24.4	21.4	32.2






#36
 Benzo(a,h,i)perylene
 Concen: 4.37 ng
 RT: 27.21 min Scan# 1752
 Delta R.T. -0.02 min
 Lab File: BE092786.D
 Acq: 6 Mar 2017 14:00

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC005

Tot Ion	Ratio	Lower	Upper
276	100		
277	24.4	19.8	29.8
138	22.5	19.8	29.6

Manual Integrations
 APPROVED

mohammad
 3/7/2017 4:22:49 PM



Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

