

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE030618\
 Data File : BE095700.D
 Acq On : 6 Mar 2018 16:46
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 07 01:10:07 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE022318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 23 12:12:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.630	0.661	-4.9	105	0.00
3	n-Nitrosodimethylamine	0.781	0.831	-6.4	106	0.00
4 S	2-Fluorophenol	1.056	1.098	-4.0	104	0.00
5 S	Phenol-d6	1.462	1.506	-3.0	100	0.00
6	bis(2-Chloroethyl)ether	1.500	1.586	-5.7	103	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
8 S	Nitrobenzene-d5	0.412	0.394	4.4	98	0.00
9	Naphthalene	4.611	4.674	-1.4	99	0.00
10	Hexachlorobutadiene	0.269	0.240	10.8	84	0.00
11 SURR	2-Methylnaphthalene-d10	0.669	0.669	0.0	99	0.00
12	2-Methylnaphthalene	0.796	0.792	0.5	99	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
14 S	2,4,6-Tribromophenol	0.143	0.125	12.6	101	0.00
15 S	2-Fluorobiphenyl	1.787	1.721	3.7	98	0.00
16	Acenaphthylene	2.247	2.090	7.0	97	0.00
17	Acenaphthene	1.411	1.364	3.3	99	0.00
18	Fluorene	1.750	1.684	3.8	100	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
20	4-Bromophenyl-phenylether	0.251	0.226	10.0	97	0.01
21	Hexachlorobenzene	0.260	0.232	10.8	97	0.00
22	Pentachlorophenol	0.060	0.052	13.3	115	-0.01
23	Phenanthrene	1.250	1.177	5.8	100	0.00
24	Anthracene	1.148	1.057	7.9	103	0.00
25 SURR	Fluoranthene-d10	5.225	4.865	6.9	102	0.00
26	Fluoranthene	1.538	1.436	6.6	101	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
28	Pvrene	1.636	1.615	1.3	105	0.00
29 S	Terphenyl-d14	0.797	0.782	1.9	103	0.00
30	Benzo(a)anthracene	1.305	1.293	0.9	106	0.00
31	Chrysene	1.298	1.236	4.8	100	0.00
32	Bis(2-ethylhexyl)phthalate	2.308	1.985	14.0	100	0.00
33	Indeno(1,2,3-cd)pyrene	1.257	1.208	3.9	102	0.00
34 I	Perylene-d12	1.000	1.000	0.0	108	0.00
35	Benzo(b)fluoranthene	1.425	1.330	6.7	99	0.00
36	Benzo(k)fluoranthene	1.417	1.369	3.4	103	0.00
37 C	Benzo(a)pyrene	1.284	1.252	2.5	104	0.00
38	Dibenzo(a,h)anthracene	1.143	1.099	3.8	104	0.00
39	Benzo(a,h,i)perylene	1.224	1.139	6.9	99	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE030618\
Data File : BE095700.D
Acq On : 6 Mar 2018 16:46
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 07 01:10:07 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE022318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 23 12:12:42 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		