

Data Path : Z:\HPCHEM1\BNA E\DATA\BE030717\
 Data File : BE092809.D
 Acq On : 7 Mar 2017 17:07
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 08 04:30:12 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE030617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 06 15:46:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	-0.02
2	1,4-Dioxane	0.743	0.804	-8.2	112	-0.01
3	n-Nitrosodimethylamine	0.839	0.600	28.5#	74	-0.01
4 S	2-Fluorophenol	1.058	0.746	29.5#	77	0.00
5 S	Phenol-d6	1.260	0.809	35.8#	71	0.02
6	bis(2-Chloroethyl)ether	1.374	1.413	-2.8	105	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	94	-0.02
8 S	Nitrobenzene-d5	0.306	0.205	33.0#	77	0.00
9	Naphthalene	1.070	1.100	-2.8	106	0.00
10	Hexachlorobutadiene	0.154	0.169	-9.7	108	-0.02
11	2-Methylnaphthalene	0.692	0.647	6.5	99	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	108	-0.02
13 S	2,4,6-Tribromophenol	0.132	0.092	30.3#	77	-0.02
14 S	2-Fluorobiphenyl	1.179	1.118	5.2	102	-0.01
15	Acenaphthylene	7.684	7.290	5.1	107	-0.02
16	Acenaphthene	1.076	1.043	3.1	104	-0.02
17	Fluorene	1.463	1.392	4.9	103	-0.01
18 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
19	Hexachlorobenzene	0.205	0.205	0.0	112	-0.02
20	Pentachlorophenol	0.069	0.046	33.3#	81	-0.09
21	Phenanthrene	1.199	1.179	1.7	107	-0.02
22	Anthracene	1.209	1.124	7.0	104	0.00
23	Fluoranthene	5.839	5.724	2.0	107	-0.02
24 I	Chrysene-d12	1.000	1.000	0.0	102	-0.02
25	Pyrene	4.529	4.339	4.2	108	0.00
26 S	Terphenyl-d14	0.603	0.557	7.6	102	-0.02
27	Benzo(a)anthracene	5.028	4.350	13.5	108	-0.02
28	Chrysene	4.585	4.906	-7.0	104	-0.02
29	Bis(2-ethylhexyl)phthalate	0.549	0.314	42.8#	83	-0.02
30	Indeno(1,2,3-cd)pyrene	4.965	4.465	10.1	107	-0.03
31 I	Perylene-d12	1.000	1.000	0.0	97	-0.02
32	Benzo(b)fluoranthene	1.332	1.304	2.1	114	-0.02
33	Benzo(k)fluoranthene	1.340	1.360	-1.5	102	-0.02
34 C	Benzo(a)pyrene	1.211	1.189	1.8	110	-0.02
35	Dibenzo(a,h)anthracene	1.157	1.052	9.1	104	-0.02
36	Benzo(g,h,i)perylene	4.951	4.715	4.8	104	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0