

Data Path : Z:\HPCHEM1\BNA E\DATA\BE030717\
 Data File : BE092809.D
 Acq On : 7 Mar 2017 17:07
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 08 04:30:12 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE030617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 06 15:46:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	90	-0.02
2	1,4-Dioxane	0.400	0.438	-9.5	112	-0.01
3	n-Nitrosodimethylamine	0.400	0.286	28.5#	74	-0.01
4 S	2-Fluorophenol	0.400	0.282	29.5#	77	0.00
5 S	Phenol-d6	0.400	0.257	35.8#	71	0.02
6	bis(2-Chloroethyl)ether	0.400	0.411	-2.7	105	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	94	-0.02
8 S	Nitrobenzene-d5	0.400	0.269	32.8#	77	0.00
9	Naphthalene	0.400	0.411	-2.7	106	0.00
10	Hexachlorobutadiene	0.400	0.438	-9.5	108	-0.02
11	2-Methylnaphthalene	0.400	0.374	6.5	99	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	108	-0.02
13 S	2,4,6-Tribromophenol	0.400	0.281	29.8#	77	-0.02
14 S	2-Fluorobiphenyl	0.400	0.379	5.3	102	-0.01
15	Acenaphthylene	0.400	0.379	5.3	107	-0.02
16	Acenaphthene	0.400	0.388	3.0	104	-0.02
17	Fluorene	0.400	0.381	4.8	103	-0.01
18 I	Phenanthrene-d10	5.000	5.000	0.0	96	0.00
19	Hexachlorobenzene	0.400	0.401	-0.3	112	-0.02
20	Pentachlorophenol	0.400	0.330	17.5	81	-0.09
21	Phenanthrene	0.400	0.393	1.8	107	-0.02
22	Anthracene	0.400	0.372	7.0	104	0.00
23	Fluoranthene	0.400	0.392	2.0	107	-0.02
24 I	Chrysene-d12	5.000	5.000	0.0	102	-0.02
25	Pyrene	0.400	0.383	4.3	108	0.00
26 S	Terphenyl-d14	0.400	0.370	7.5	102	-0.02
27	Benzo(a)anthracene	0.400	0.397	0.8	108	-0.02
28	Chrysene	0.400	0.428	-7.0	104	-0.02
29	Bis(2-ethylhexyl)phthalate	0.400	0.301	24.8	83	-0.02
30	Indeno(1,2,3-cd)pyrene	0.400	0.360	10.0	107	-0.03
31 I	Perylene-d12	5.000	5.000	0.0	97	-0.02
32	Benzo(b)fluoranthene	0.400	0.392	2.0	114	-0.02
33	Benzo(k)fluoranthene	0.400	0.406	-1.5	102	-0.02
34 C	Benzo(a)pyrene	0.400	0.393	1.8	110	-0.02
35	Dibenzo(a,h)anthracene	0.400	0.364	9.0	104	-0.02
36	Benzo(g,h,i)perylene	0.400	0.381	4.8	104	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0