

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE030718\
 Data File : BE095718.D
 Acq On : 7 Mar 2018 11:40
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 07 14:22:01 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE022318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 23 12:12:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.630	0.618	1.9	101	0.00
3	n-Nitrosodimethylamine	0.781	0.788	-0.9	104	0.00
4 S	2-Fluorophenol	1.056	1.166	-10.4	114	0.00
5 S	Phenol-d6	1.462	1.626	-11.2	112	0.00
6	bis(2-Chloroethyl)ether	1.500	1.575	-5.0	106	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
8 S	Nitrobenzene-d5	0.412	0.401	2.7	108	0.00
9	Naphthalene	4.611	4.660	-1.1	107	0.00
10	Hexachlorobutadiene	0.269	0.304	-13.0	115	0.00
11 SURR	2-Methylnaphthalene-d10	0.669	0.675	-0.9	108	0.00
12	2-Methylnaphthalene	0.796	0.802	-0.8	108	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.00
14 S	2,4,6-Tribromophenol	0.143	0.133	7.0	119	0.00
15 S	2-Fluorobiphenyl	1.787	1.702	4.8	107	0.00
16	Acenaphthylene	2.247	2.107	6.2	108	0.00
17	Acenaphthene	1.411	1.337	5.2	108	0.00
18	Fluorene	1.750	1.641	6.2	107	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	114	0.00
20	4-Bromophenyl-phenylether	0.251	0.237	5.6	105	0.01
21	Hexachlorobenzene	0.260	0.244	6.2	105	0.00
22	Pentachlorophenol	0.060	0.060	0.0	139	0.00
23	Phenanthrene	1.250	1.169	6.5	103	0.00
24	Anthracene	1.148	1.091	5.0	110	0.00
25 SURR	Fluoranthene-d10	5.225	4.823	7.7	104	0.00
26	Fluoranthene	1.538	1.413	8.1	103	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
28	Pvrene	1.636	1.678	-2.6	110	0.00
29 S	Terphenyl-d14	0.797	0.792	0.6	106	-0.01
30	Benzo(a)anthracene	1.305	1.317	-0.9	109	0.00
31	Chrysene	1.298	1.264	2.6	104	0.00
32	Bis(2-ethylhexyl)phthalate	2.308	2.365	-2.5	121	0.00
33	Indeno(1,2,3-cd)pyrene	1.257	1.323	-5.3	113	0.00
34 I	Perylene-d12	1.000	1.000	0.0	115	0.00
35	Benzo(b)fluoranthene	1.425	1.319	7.4	104	0.00
36	Benzo(k)fluoranthene	1.417	1.359	4.1	109	0.00
37 C	Benzo(a)pyrene	1.284	1.218	5.1	107	0.00
38	Dibenzo(a,h)anthracene	1.143	1.135	0.7	114	0.00
39	Benzo(a,h,i)perylene	1.224	1.160	5.2	107	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE030718\
Data File : BE095718.D
Acq On : 7 Mar 2018 11:40
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 07 14:22:01 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE022318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 23 12:12:42 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			