

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE030719\
 Data File : BE098827.D
 Acq On : 7 Mar 2019 19:15
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 08 04:01:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 15:33:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	138	0.00
2	1,4-Dioxane	0.789	0.635	19.5	115	0.00
3	n-Nitrosodimethylamine	1.006	0.824	18.1	111	0.00
4 S	2-Fluorophenol	1.196	1.072	10.4	119	0.00
5 S	Phenol-d6	1.689	1.656	2.0	135	-0.01
6	bis(2-Chloroethyl)ether	1.304	1.347	-3.3	148	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	142	0.00
8 S	Nitrobenzene-d5	0.419	0.384	8.4	145	-0.01
9	Naphthalene	4.581	4.494	1.9	143	-0.01
10	Hexachlorobutadiene	0.303	0.291	4.0	141	0.00
11 SURR	2-Methylnaphthalene-d10	0.739	0.723	2.2	142	0.00
12	2-Methylnaphthalene	0.743	0.733	1.3	145	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	147	0.00
14 S	2,4,6-Tribromophenol	0.202	0.196	3.0	144	-0.01
15 S	2-Fluorobiphenyl	1.656	1.642	0.8	148	-0.01
16	Acenaphthylene	2.057	2.058	-0.0	154#	0.00
17	Acenaphthene	1.220	1.211	0.7	147	-0.01
18	Fluorene	1.760	1.691	3.9	145	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	140	-0.01
20	4-Bromophenyl-phenylether	0.217	0.211	2.8	139	0.00
21	Hexachlorobenzene	0.258	0.261	-1.2	149	0.00
22	Pentachlorophenol	0.052	0.047	9.6	172#	-0.01
23	Phenanthrene	1.137	1.131	0.5	142	-0.01
24	Anthracene	0.972	0.911	6.3	138	0.00
25 SURR	Fluoranthene-d10	6.385	5.898	7.6	134	0.00
26	Fluoranthene	1.605	1.484	7.5	134	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	134	-0.01
28	Pvrene	1.224	1.225	-0.1	134	0.00
29 S	Terphenyl-d14	0.972	0.965	0.7	134	0.00
30	Benzo(a)anthracene	1.248	1.195	4.2	131	0.00
31	Chrysene	1.336	1.320	1.2	136	0.00
32	Bis(2-ethylhexyl)phthalate	2.657	2.842	-7.0	156#	0.00
33	Indeno(1,2,3-cd)pyrene	1.410	1.429	-1.3	138	0.00
34 I	Perylene-d12	1.000	1.000	0.0	139	0.00
35	Benzo(b)fluoranthene	1.315	1.326	-0.8	145	0.00
36	Benzo(k)fluoranthene	1.377	1.384	-0.5	146	0.00
37 C	Benzo(a)pyrene	1.261	1.247	1.1	142	0.00
38	Dibenzo(a,h)anthracene	1.193	1.176	1.4	139	-0.01
39	Benzo(a,h,i)perylene	1.248	1.198	4.0	136	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE030719\
Data File : BE098827.D
Acq On : 7 Mar 2019 19:15
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 08 04:01:25 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 06 15:33:24 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			