

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE030719\
 Data File : BE098827.D
 Acq On : 7 Mar 2019 19:15
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 08 04:01:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 15:33:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	138	0.00
2	1,4-Dioxane	0.400	0.308	23.0	115	0.00
3	n-Nitrosodimethylamine	0.400	0.328	18.0	111	0.00
4 S	2-Fluorophenol	0.400	0.359	10.3	119	0.00
5 S	Phenol-d6	0.400	0.392	2.0	135	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.413	-3.2	148	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	142	0.00
8 S	Nitrobenzene-d5	0.400	0.367	8.3	145	-0.01
9	Naphthalene	0.400	0.392	2.0	143	-0.01
10	Hexachlorobutadiene	0.400	0.384	4.0	141	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.391	2.3	142	0.00
12	2-Methylnaphthalene	0.400	0.395	1.3	145	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	147	0.00
14 S	2,4,6-Tribromophenol	0.400	0.389	2.8	144	-0.01
15 S	2-Fluorobiphenyl	0.400	0.397	0.8	148	-0.01
16	Acenaphthylene	0.400	0.400	0.0	154	0.00
17	Acenaphthene	0.400	0.397	0.8	147	-0.01
18	Fluorene	0.400	0.384	4.0	145	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	140	-0.01
20	4-Bromophenyl-phenylether	0.400	0.390	2.5	139	0.00
21	Hexachlorobenzene	0.400	0.404	-1.0	149	0.00
22	Pentachlorophenol	0.400	0.481	-20.2	172	-0.01
23	Phenanthrene	0.400	0.398	0.5	142	-0.01
24	Anthracene	0.400	0.375	6.3	138	0.00
25 SURR	Fluoranthene-d10	0.400	0.369	7.8	134	0.00
26	Fluoranthene	0.400	0.370	7.5	134	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	134	-0.01
28	Pvrene	0.400	0.400	0.0	134	0.00
29 S	Terphenyl-d14	0.400	0.397	0.8	134	0.00
30	Benzo(a)anthracene	0.400	0.383	4.3	131	0.00
31	Chrysene	0.400	0.395	1.3	136	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.428	-7.0	156	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.405	-1.3	138	0.00
34 I	Perylene-d12	0.400	0.400	0.0	139	0.00
35	Benzo(b)fluoranthene	0.400	0.403	-0.8	145	0.00
36	Benzo(k)fluoranthene	0.400	0.402	-0.5	146	0.00
37 C	Benzo(a)pyrene	0.400	0.395	1.3	142	0.00
38	Dibenzo(a,h)anthracene	0.400	0.394	1.5	139	-0.01
39	Benzo(a,h,i)perylene	0.400	0.384	4.0	136	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE030719\
Data File : BE098827.D
Acq On : 7 Mar 2019 19:15
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 08 04:01:25 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 06 15:33:24 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				