

Data Path : Z:\HPCHEM1\BNA_E\Data\BE031017\
 Data File : BE092811.D
 Acq On : 10 Mar 2017 17:22
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 10 22:58:49 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE030617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 06 15:46:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	103	-0.02
2	1,4-Dioxane	0.400	0.381	4.8	114	-0.01
3	n-Nitrosodimethylamine	0.400	0.284	29.0#	84	-0.01
4 S	2-Fluorophenol	0.400	0.288	28.0#	89	0.00
5 S	Phenol-d6	0.400	0.275	31.3#	86	0.00
6	bis(2-Chloroethyl)ether	0.400	0.328	18.0	96	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	110	-0.02
8 S	Nitrobenzene-d5	0.400	0.318	20.5	107	-0.02
9	Naphthalene	0.400	0.417	-4.2	126	0.00
10	Hexachlorobutadiene	0.400	0.432	-8.0	125	-0.02
11	2-Methylnaphthalene	0.400	0.384	4.0	119	-0.01
12 I	Acenaphthene-d10	5.000	5.000	0.0	126	-0.02
13 S	2,4,6-Tribromophenol	0.400	0.278	30.5#	89	0.00
14 S	2-Fluorobiphenyl	0.400	0.394	1.5	124	-0.01
15	Acenaphthylene	0.400	0.385	3.8	126	-0.02
16	Acenaphthene	0.400	0.399	0.3	125	-0.02
17	Fluorene	0.400	0.389	2.8	122	-0.01
18 I	Phenanthrene-d10	5.000	5.000	0.0	110	0.00
19	Hexachlorobenzene	0.400	0.408	-2.0	131	-0.02
20	Pentachlorophenol	0.400	0.227	43.3#	54	-0.05
21	Phenanthrene	0.400	0.398	0.5	124	0.00
22	Anthracene	0.400	0.387	3.3	124	0.00
23	Fluoranthene	0.400	0.379	5.3	119	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	99	-0.02
25	Pyrene	0.400	0.430	-7.5	118	0.00
26 S	Terphenyl-d14	0.400	0.426	-6.5	115	-0.02
27	Benzo(a)anthracene	0.400	0.555	-38.8#	156	-0.02
28	Chrysene	0.400	0.322	19.5	77	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.404	-1.0	105	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.382	4.5	111	-0.02
31 I	Perylene-d12	5.000	5.000	0.0	103	-0.01
32	Benzo(b)fluoranthene	0.400	0.410	-2.5	127	-0.01
33	Benzo(k)fluoranthene	0.400	0.382	4.5	102	-0.01
34 C	Benzo(a)pyrene	0.400	0.387	3.3	115	-0.02
35	Dibenzo(a,h)anthracene	0.400	0.356	11.0	108	-0.02
36	Benzo(g,h,i)perylene	0.400	0.379	5.3	110	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0