

Data Path : Z:\HPCHEM1\BNA_E\Data\BE031017\
 Data File : BE092816.D
 Acq On : 10 Mar 2017 20:30
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 10 23:05:46 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE030617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 06 15:46:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	98	-0.02
2	1,4-Dioxane	0.400	0.382	4.5	110	-0.01
3	n-Nitrosodimethylamine	0.400	0.258	35.5#	73	0.00
4 S	2-Fluorophenol	0.400	0.272	32.0#	81	0.00
5 S	Phenol-d6	0.400	0.242	39.5#	73	0.02
6	bis(2-Chloroethyl)ether	0.400	0.423	-5.7	118	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	102	-0.02
8 S	Nitrobenzene-d5	0.400	0.298	25.5#	93	0.00
9	Naphthalene	0.400	0.414	-3.5	116	0.00
10	Hexachlorobutadiene	0.400	0.428	-7.0	115	-0.02
11	2-Methylnaphthalene	0.400	0.377	5.8	109	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	113	-0.02
13 S	2,4,6-Tribromophenol	0.400	0.276	31.0#	79	-0.01
14 S	2-Fluorobiphenyl	0.400	0.391	2.3	110	-0.01
15	Acenaphthylene	0.400	0.383	4.3	113	0.00
16	Acenaphthene	0.400	0.397	0.8	112	-0.02
17	Fluorene	0.400	0.379	5.3	107	-0.01
18 I	Phenanthrene-d10	5.000	5.000	0.0	102	0.00
19	Hexachlorobenzene	0.400	0.373	6.8	111	-0.02
20	Pentachlorophenol	0.400	0.213	46.8#	45	-0.09
21	Phenanthrene	0.400	0.372	7.0	108	0.00
22	Anthracene	0.400	0.366	8.5	110	0.00
23	Fluoranthene	0.400	0.368	8.0	107	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	93	-0.02
25	Pyrene	0.400	0.409	-2.2	106	0.00
26 S	Terphenyl-d14	0.400	0.403	-0.8	102	0.00
27	Benzo(a)anthracene	0.400	0.537	-34.3#	141	-0.02
28	Chrysene	0.400	0.341	14.8	76	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.337	15.8	84	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.404	-1.0	110	-0.02
31 I	Perylene-d12	5.000	5.000	0.0	100	-0.02
32	Benzo(b)fluoranthene	0.400	0.366	8.5	109	-0.02
33	Benzo(k)fluoranthene	0.400	0.426	-6.5	110	-0.02
34 C	Benzo(a)pyrene	0.400	0.401	-0.3	115	-0.02
35	Dibenzo(a,h)anthracene	0.400	0.364	9.0	107	0.00
36	Benzo(g,h,i)perylene	0.400	0.387	3.3	108	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0