

Data Path : Z:\HPCHEM1\BNA_E\Data\BE031017\
 Data File : BE092816.D
 Acq On : 10 Mar 2017 20:30
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 10 23:05:46 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE030617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 06 15:46:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	-0.02
2	1,4-Dioxane	0.743	0.723	2.7	110	-0.01
3	n-Nitrosodimethylamine	0.839	0.542	35.4#	73	0.00
4 S	2-Fluorophenol	1.058	0.719	32.0#	81	0.00
5 S	Phenol-d6	1.260	0.764	39.4#	73	0.02
6	bis(2-Chloroethyl)ether	1.374	1.455	-5.9	118	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	102	-0.02
8 S	Nitrobenzene-d5	0.306	0.228	25.5#	93	0.00
9	Naphthalene	1.070	1.106	-3.4	116	0.00
10	Hexachlorobutadiene	0.154	0.165	-7.1	115	-0.02
11	2-Methylnaphthalene	0.692	0.651	5.9	109	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	113	-0.02
13 S	2,4,6-Tribromophenol	0.132	0.091	31.1#	79	-0.01
14 S	2-Fluorobiphenyl	1.179	1.153	2.2	110	-0.01
15	Acenaphthylene	7.684	7.362	4.2	113	0.00
16	Acenaphthene	1.076	1.067	0.8	112	-0.02
17	Fluorene	1.463	1.386	5.3	107	-0.01
18 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
19	Hexachlorobenzene	0.205	0.191	6.8	111	-0.02
20	Pentachlorophenol	0.069	0.024	65.2#	45#	-0.09
21	Phenanthrene	1.199	1.116	6.9	108	0.00
22	Anthracene	1.209	1.105	8.6	110	0.00
23	Fluoranthene	5.839	5.365	8.1	107	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	93	-0.02
25	Pyrene	4.529	4.632	-2.3	106	0.00
26 S	Terphenyl-d14	0.603	0.608	-0.8	102	0.00
27	Benzo(a)anthracene	5.028	6.189	-23.1	141	-0.02
28	Chrysene	4.585	3.908	14.8	76	0.00
29	Bis(2-ethylhexyl)phthalate	0.549	0.347	36.8#	84	0.00
30	Indeno(1,2,3-cd)pyrene	4.965	5.020	-1.1	110	-0.02
31 I	Perylene-d12	1.000	1.000	0.0	100	-0.02
32	Benzo(b)fluoranthene	1.332	1.219	8.5	109	-0.02
33	Benzo(k)fluoranthene	1.340	1.426	-6.4	110	-0.02
34 C	Benzo(a)pyrene	1.211	1.214	-0.2	115	-0.02
35	Dibenzo(a,h)anthracene	1.157	1.052	9.1	107	0.00
36	Benzo(g,h,i)perylene	4.951	4.787	3.3	108	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0