

Data Path : Z:\HPCHEM1\BNA E\DATA\BE031218\  
 Data File : BE095725.D  
 Acq On : 12 Mar 2018 12:02  
 Operator : SJ/JU  
 Sample : J1867-01  
 Misc :  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 ClientSampleId :  
 CPS133071

Quant Time: Mar 12 12:52:19 2018  
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE022318.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Feb 23 12:12:42 2018  
 Response via : Initial Calibration

| Internal Standards             | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|--------------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4      | 8.43  | 152  | 1525     | 0.40  | ng    | 0.00     |
| 7) Naphthalene-d8              | 11.27 | 136  | 6091     | 0.40  | ng    | 0.00     |
| 13) Acenaphthene-d10           | 15.00 | 164  | 3440     | 0.40  | ng    | 0.00     |
| 19) Phenanthrene-d10           | 17.71 | 188  | 7730     | 0.40  | ng    | 0.00     |
| 27) Chrysene-d12               | 21.80 | 240  | 7818     | 0.40  | ng    | 0.01     |
| 34) Perylene-d12               | 24.43 | 264  | 7291     | 0.40  | ng    | 0.00     |
| System Monitoring Compounds    |       |      |          |       |       |          |
| 4) 2-Fluorophenol              | 5.92  | 112  | 760      | 0.19  | ng    | 0.00     |
| 5) Phenol-d6                   | 7.56  | 99   | 680      | 0.12  | ng    | 0.00     |
| 8) Nitrobenzene-d5             | 9.61  | 82   | 2375     | 0.38  | ng    | 0.00     |
| 11) 2-Methylnaphthalene-d10    | 12.81 | 152  | 4045     | 0.40  | ng    | 0.00     |
| 14) 2,4,6-Tribromophenol       | 16.47 | 330  | 545      | 0.45  | ng    | 0.00     |
| 15) 2-Fluorobiphenyl           | 13.64 | 172  | 7311     | 0.48  | ng    | 0.00     |
| 25) Fluoranthene-d10           | 19.69 | 212  | 41616    | 0.41  | ng    | 0.00     |
| 29) Terphenyl-d14              | 20.24 | 244  | 7976     | 0.51  | ng    | 0.00     |
| Target Compounds               |       |      |          |       |       | Qvalue   |
| 32) Bis(2-ethylhexyl)phthalate | 21.66 | 149  | 4183     | 0.093 | ng    | 98       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE031218\  
Data File : BE095725.D  
Acq On : 12 Mar 2018 12:02  
Operator : SJ/JU  
Sample : J1867-01  
Misc :  
ALS Vial : 6 Sample Multiplier: 1

Instrument :  
BNA\_E  
ClientSampleId :  
CPS133071

Quant Time: Mar 12 12:52:19 2018  
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE022318.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Feb 23 12:12:42 2018  
Response via : Initial Calibration

