

Data Path : Z:\HPCHEM1\BNA E\DATA\BE031218\
 Data File : BE095727.D
 Acq On : 12 Mar 2018 13:14
 Operator : SJ/JU
 Sample : J1867-03
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 CPS093071

Quant Time: Mar 12 14:42:42 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE022318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 23 12:12:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.43	152	1678	0.40	ng	0.00
7) Naphthalene-d8	11.27	136	6588	0.40	ng	0.00
13) Acenaphthene-d10	15.00	164	4608	0.40	ng	0.00
19) Phenanthrene-d10	17.71	188	8631	0.40	ng	0.00
27) Chrysene-d12	21.80	240	8722	0.40	ng	0.00
34) Perylene-d12	24.43	264	8194	0.40	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.92	112	776	0.18	ng	0.00
5) Phenol-d6	7.56	99	727	0.12	ng	0.00
8) Nitrobenzene-d5	9.61	82	2697	0.40	ng	0.00
11) 2-Methylnaphthalene-d10	12.80	152	4079	0.37	ng	0.00
14) 2,4,6-Tribromophenol	16.47	330	736	0.45	ng	0.00
15) 2-Fluorobiphenyl	13.64	172	6909	0.34	ng	0.00
25) Fluoranthene-d10	19.69	212	42669	0.38	ng	0.00
29) Terphenyl-d14	20.25	244	6853	0.39	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) Naphthalene	11.31	128	44727	0.589	ng	# 96
12) 2-Methylnaphthalene	12.87	142	31273	2.387	ng	# 97
17) Acenaphthene	15.06	154	2163	0.133	ng	# 56
18) Fluorene	16.03	166	4532	0.225	ng	# 81
22) Pentachlorophenol	17.37	266	360391	175.119	ng	# 71
23) Phenanthrene	17.75	178	1481	0.055	ng	# 96
32) Bis(2-ethylhexyl)phthalate	21.66	149	7786	0.155	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE031218\
Data File : BE095727.D
Acq On : 12 Mar 2018 13:14
Operator : SJ/JU
Sample : J1867-03
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
CPS093071

Quant Time: Mar 12 14:42:42 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE022318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 23 12:12:42 2018
Response via : Initial Calibration

