

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE031218\  
 Data File : BE095723.D  
 Acq On : 12 Mar 2018 9:39  
 Operator : SJ/JU  
 Sample : SSTDCCC0.4  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 LabSampleId :  
 SSTDCCC0.4

Quant Time: Mar 12 12:24:23 2018  
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE022318.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Feb 23 12:12:42 2018  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|         | Compound                   | AvqRF | CCRF  | %Dev | Area% | Dev(min) |
|---------|----------------------------|-------|-------|------|-------|----------|
| 1 I     | 1,4-Dichlorobenzene-d4     | 1.000 | 1.000 | 0.0  | 116   | 0.00     |
| 2       | 1,4-Dioxane                | 0.630 | 0.538 | 14.6 | 102   | 0.00     |
| 3       | n-Nitrosodimethylamine     | 0.781 | 0.722 | 7.6  | 111   | 0.00     |
| 4 S     | 2-Fluorophenol             | 1.056 | 0.940 | 11.0 | 107   | 0.00     |
| 5 S     | Phenol-d6                  | 1.462 | 1.407 | 3.8  | 112   | 0.00     |
| 6       | bis(2-Chloroethyl)ether    | 1.500 | 1.433 | 4.5  | 111   | 0.00     |
| 7 I     | Naphthalene-d8             | 1.000 | 1.000 | 0.0  | 122   | 0.00     |
| 8 S     | Nitrobenzene-d5            | 0.412 | 0.370 | 10.2 | 112   | 0.00     |
| 9       | Naphthalene                | 4.611 | 4.193 | 9.1  | 108   | 0.00     |
| 10      | Hexachlorobutadiene        | 0.269 | 0.267 | 0.7  | 114   | 0.00     |
| 11 SURR | 2-Methylnaphthalene-d10    | 0.669 | 0.600 | 10.3 | 108   | 0.00     |
| 12      | 2-Methylnaphthalene        | 0.796 | 0.713 | 10.4 | 108   | 0.00     |
| 13 I    | Acenaphthene-d10           | 1.000 | 1.000 | 0.0  | 124   | 0.00     |
| 14 S    | 2,4,6-Tribromophenol       | 0.143 | 0.132 | 7.7  | 128   | 0.00     |
| 15 S    | 2-Fluorobiphenyl           | 1.787 | 1.572 | 12.0 | 107   | 0.00     |
| 16      | Acenaphthylene             | 2.247 | 1.999 | 11.0 | 111   | 0.00     |
| 17      | Acenaphthene               | 1.411 | 1.256 | 11.0 | 109   | 0.00     |
| 18      | Fluorene                   | 1.750 | 1.567 | 10.5 | 111   | 0.00     |
| 19 I    | Phenanthrene-d10           | 1.000 | 1.000 | 0.0  | 129   | 0.00     |
| 20      | 4-Bromophenyl-phenylether  | 0.251 | 0.219 | 12.7 | 110   | 0.00     |
| 21      | Hexachlorobenzene          | 0.260 | 0.224 | 13.8 | 110   | 0.00     |
| 22      | Pentachlorophenol          | 0.060 | 0.053 | 11.7 | 139   | -0.01    |
| 23      | Phenanthrene               | 1.250 | 1.116 | 10.7 | 112   | 0.00     |
| 24      | Anthracene                 | 1.148 | 1.018 | 11.3 | 117   | 0.00     |
| 25 SURR | Fluoranthene-d10           | 5.225 | 4.729 | 9.5  | 117   | 0.00     |
| 26      | Fluoranthene               | 1.538 | 1.388 | 9.8  | 115   | 0.00     |
| 27 I    | Chrysene-d12               | 1.000 | 1.000 | 0.0  | 131   | 0.00     |
| 28      | Pvrene                     | 1.636 | 1.411 | 13.8 | 113   | 0.00     |
| 29 S    | Terphenyl-d14              | 0.797 | 0.723 | 9.3  | 118   | -0.01    |
| 30      | Benzo(a)anthracene         | 1.305 | 1.242 | 4.8  | 125   | 0.00     |
| 31      | Chrysene                   | 1.298 | 1.141 | 12.1 | 114   | 0.00     |
| 32      | Bis(2-ethylhexyl)phthalate | 2.308 | 2.269 | 1.7  | 141   | 0.00     |
| 33      | Indeno(1,2,3-cd)pyrene     | 1.257 | 1.139 | 9.4  | 118   | 0.00     |
| 34 I    | Perylene-d12               | 1.000 | 1.000 | 0.0  | 135   | 0.00     |
| 35      | Benzo(b)fluoranthene       | 1.425 | 1.230 | 13.7 | 114   | 0.00     |
| 36      | Benzo(k)fluoranthene       | 1.417 | 1.259 | 11.2 | 119   | 0.00     |
| 37 C    | Benzo(a)pyrene             | 1.284 | 1.133 | 11.8 | 118   | 0.00     |
| 38      | Dibenzo(a,h)anthracene     | 1.143 | 1.026 | 10.2 | 122   | 0.00     |
| 39      | Benzo(a,h,i)perylene       | 1.224 | 1.070 | 12.6 | 117   | 0.00     |

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE031218\  
Data File : BE095723.D  
Acq On : 12 Mar 2018 9:39  
Operator : SJ/JU  
Sample : SSTDCCC0.4  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_E  
LabSampleId :  
SSTDCCC0.4

Quant Time: Mar 12 12:24:23 2018  
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE022318.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Feb 23 12:12:42 2018  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound           | AvgRF                        | CCRF | %Dev Area | % Dev(min) |
|--------------------|------------------------------|------|-----------|------------|
| -----              |                              |      |           |            |
| (#) = Out of Range | SPCC's out = 0 CCC's out = 0 |      |           |            |