

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE031218\  
 Data File : BE095738.D  
 Acq On : 12 Mar 2018 19:54  
 Operator : SJ/JU  
 Sample : SSTDCCC0.4  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 LabSampleId :  
 SSTDCCC0.4

Quant Time: Mar 13 10:48:05 2018  
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE022318.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Feb 23 12:12:42 2018  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	126	0.00
2	1,4-Dioxane	0.400	0.354	11.5	115	0.00
3	n-Nitrosodimethylamine	0.400	0.382	4.5	125	0.00
4 S	2-Fluorophenol	0.400	0.388	3.0	127	0.00
5 S	Phenol-d6	0.400	0.430	-7.5	137	0.00
6	bis(2-Chloroethyl)ether	0.400	0.371	7.3	118	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	136	-0.02
8 S	Nitrobenzene-d5	0.400	0.375	6.3	130	0.00
9	Naphthalene	0.400	0.364	9.0	120	0.00
10	Hexachlorobutadiene	0.400	0.394	1.5	126	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.367	8.3	123	0.00
12	2-Methylnaphthalene	0.400	0.365	8.8	122	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	141	0.00
14 S	2,4,6-Tribromophenol	0.400	0.445	-11.2	174	0.00
15 S	2-Fluorobiphenyl	0.400	0.353	11.8	122	0.00
16	Acenaphthylene	0.400	0.361	9.8	128	-0.01
17	Acenaphthene	0.400	0.359	10.3	125	0.00
18	Fluorene	0.400	0.364	9.0	128	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	138	0.00
20	4-Bromophenyl-phenylether	0.400	0.356	11.0	120	0.01
21	Hexachlorobenzene	0.400	0.353	11.8	120	0.00
22	Pentachlorophenol	0.400	0.534	-33.5#	238	-0.01
23	Phenanthrene	0.400	0.360	10.0	120	0.00
24	Anthracene	0.400	0.361	9.8	127	0.00
25 SURR	Fluoranthene-d10	0.400	0.358	10.5	123	0.00
26	Fluoranthene	0.400	0.350	12.5	119	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	134	0.00
28	Pyrene	0.400	0.376	6.0	126	0.00
29 S	Terphenyl-d14	0.400	0.371	7.3	124	-0.01
30	Benzo(a)anthracene	0.400	0.377	5.8	127	0.00
31	Chrysene	0.400	0.356	11.0	119	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.469	-17.2	173	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.385	3.8	129	0.00
34 I	Perylene-d12	0.400	0.400	0.0	143	0.00
35	Benzo(b)fluoranthene	0.400	0.341	14.8	119	0.00
36	Benzo(k)fluoranthene	0.400	0.350	12.5	124	0.00
37 C	Benzo(a)pyrene	0.400	0.355	11.3	125	0.00
38	Dibenzo(a,h)anthracene	0.400	0.374	6.5	134	0.00
39	Benzo(a,h,i)perylene	0.400	0.353	11.8	124	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE031218\  
Data File : BE095738.D  
Acq On : 12 Mar 2018 19:54  
Operator : SJ/JU  
Sample : SSTDCCC0.4  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_E  
LabSampleId :  
SSTDCCC0.4

Quant Time: Mar 13 10:48:05 2018  
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE022318.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Feb 23 12:12:42 2018  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
-----					
(#) = Out of Range	SPCC's out = 0		CCC's out = 0		