

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE031218\
 Data File : BE095745.D
 Acq On : 13 Mar 2018 1:19
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 13 10:50:33 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE022318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 23 12:12:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	133	0.00
2	1,4-Dioxane	0.630	0.544	13.7	119	0.00
3	n-Nitrosodimethylamine	0.781	0.724	7.3	128	0.00
4 S	2-Fluorophenol	1.056	1.002	5.1	131	0.00
5 S	Phenol-d6	1.462	1.492	-2.1	137	0.00
6	bis(2-Chloroethyl)ether	1.500	1.381	7.9	123	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	140	-0.02
8 S	Nitrobenzene-d5	0.412	0.369	10.4	128	0.00
9	Naphthalene	4.611	4.172	9.5	123	0.00
10	Hexachlorobutadiene	0.269	0.297	-10.4	145	0.00
11 SURR	2-Methylnaphthalene-d10	0.669	0.610	8.8	126	0.00
12	2-Methylnaphthalene	0.796	0.724	9.0	126	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	143	0.00
14 S	2,4,6-Tribromophenol	0.143	0.145	-1.4	161#	0.00
15 S	2-Fluorobiphenyl	1.787	1.619	9.4	127	0.00
16	Acenaphthylene	2.247	2.016	10.3	129	0.00
17	Acenaphthene	1.411	1.264	10.4	126	0.00
18	Fluorene	1.750	1.575	10.0	128	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	142	0.00
20	4-Bromophenyl-phenylether	0.251	0.227	9.6	126	0.00
21	Hexachlorobenzene	0.260	0.225	13.5	122	0.00
22	Pentachlorophenol	0.060	0.084	-40.0#	244#	-0.01
23	Phenanthrene	1.250	1.122	10.2	124	0.00
24	Anthracene	1.148	1.007	12.3	127	0.00
25 SURR	Fluoranthene-d10	5.225	4.565	12.6	124	0.00
26	Fluoranthene	1.538	1.323	14.0	121	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	138	0.00
28	Pvrene	1.636	1.556	4.9	131	0.00
29 S	Terphenyl-d14	0.797	0.725	9.0	124	-0.01
30	Benzo(a)anthracene	1.305	1.226	6.1	130	0.00
31	Chrysene	1.298	1.142	12.0	120	0.00
32	Bis(2-ethylhexyl)phthalate	2.308	2.327	-0.8	153#	0.00
33	Indeno(1,2,3-cd)pyrene	1.257	1.188	5.5	130	0.00
34 I	Perylene-d12	1.000	1.000	0.0	143	0.00
35	Benzo(b)fluoranthene	1.425	1.252	12.1	123	0.00
36	Benzo(k)fluoranthene	1.417	1.242	12.4	124	0.00
37 C	Benzo(a)pyrene	1.284	1.140	11.2	125	0.00
38	Dibenzo(a,h)anthracene	1.143	1.061	7.2	134	0.00
39	Benzo(a,h,i)perylene	1.224	1.084	11.4	125	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE031218\
Data File : BE095745.D
Acq On : 13 Mar 2018 1:19
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 13 10:50:33 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE022318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 23 12:12:42 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			