

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031319\
 Data File : BE098918.D
 Acq On : 13 Mar 2019 19:03
 Operator : JU/SJ
 Sample : K1856-06
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 BPS1-TT-MW301D-20190310

Quant Time: Mar 14 07:18:34 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 15:33:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	801	0.40	ng	0.00
7) Naphthalene-d8	10.60	136	3593	0.40	ng	0.01
13) Acenaphthene-d10	14.46	164	2635	0.40	ng	0.00
19) Phenanthrene-d10	17.22	188	6605	0.40	ng	0.00
27) Chrysene-d12	21.39	240	7863	0.40	ng	-0.01
34) Perylene-d12	23.90	264	7694	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.40	112	350	0.15	ng	0.01
5) Phenol-d6	7.01	99	212	0.06	ng	0.03
8) Nitrobenzene-d5	9.00	82	1256	0.33	ng	0.04
11) 2-Methylnaphthalene-d10	12.19	152	2274	0.34	ng	0.00
14) 2,4,6-Tribromophenol	15.96	330	553	0.42	ng	-0.01
15) 2-Fluorobiphenyl	13.07	172	7891	0.72	ng	-0.01
25) Fluoranthene-d10	19.24	212	34709	0.33	ng	0.00
29) Terphenyl-d14	19.83	244	10719	0.56	ng	-0.01
Target Compounds						
2) 1,4-Dioxane	3.29	88	332	0.165	ng	# 57
32) Bis(2-ethylhexyl)phthalate	21.29	149	2487	0.048	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031319\
Data File : BE098918.D
Acq On : 13 Mar 2019 19:03
Operator : JU/SJ
Sample : K1856-06
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
BPS1-TT-MW301D-20190310

Quant Time: Mar 14 07:18:34 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 06 15:33:24 2019
Response via : Initial Calibration

