

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031319\
 Data File : BE098931.D
 Acq On : 14 Mar 2019 4:21
 Operator : JU/SJ
 Sample : K1860-06
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 RE120D3-20190311

Quant Time: Mar 14 07:38:54 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 15:33:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	758	0.40	ng	0.00
7) Naphthalene-d8	10.60	136	3410	0.40	ng	0.01
13) Acenaphthene-d10	14.46	164	2448	0.40	ng	0.00
19) Phenanthrene-d10	17.21	188	5964	0.40	ng	0.00
27) Chrysene-d12	21.40	240	7178	0.40	ng	0.00
34) Perylene-d12	23.90	264	6868	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.41	112	299	0.13	ng	0.02
5) Phenol-d6	7.06	99	125	0.04	ng	0.07
8) Nitrobenzene-d5	9.03	82	1058	0.30	ng	0.07
11) 2-Methylnaphthalene-d10	12.21	152	2288	0.36	ng	0.01
14) 2,4,6-Tribromophenol	15.97	330	526	0.43	ng	0.00
15) 2-Fluorobiphenyl	13.09	172	4721	0.47	ng	0.00
25) Fluoranthene-d10	19.24	212	31937	0.34	ng	0.00
29) Terphenyl-d14	19.84	244	9956	0.57	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.29	88	291	0.145	ng	# 36
32) Bis(2-ethylhexyl)phthalate	21.29	149	1010	0.021	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031319\
Data File : BE098931.D
Acq On : 14 Mar 2019 4:21
Operator : JU/SJ
Sample : K1860-06
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
RE120D3-20190311

Quant Time: Mar 14 07:38:54 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 06 15:33:24 2019
Response via : Initial Calibration

