




Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE031319\
Data File : BE098908.D
Acq On : 13 Mar 2019 11:37
Operator : JU/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 15 18:50:17 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 15 05:31:33 2019
Response via : Initial Calibration



(22) Pentachlorophenol

16.859min (-0.035) 7951.88ng

response 15703

Ion	Exp%	Act%
266.00	100	100
264.00	83.30	59.43#
268.00	87.30	56.83#
0.00	0.00	0.00