

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031319\
 Data File : BE098923.D
 Acq On : 13 Mar 2019 22:08
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 14 07:24:44 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 15:33:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
2	1,4-Dioxane	0.789	0.894	-13.3	130	0.00
3	n-Nitrosodimethylamine	1.006	0.671	33.3#	72	0.01
4 S	2-Fluorophenol	1.196	1.136	5.0	101	0.00
5 S	Phenol-d6	1.689	1.661	1.7	109	0.00
6	bis(2-Chloroethyl)ether	1.304	1.119	14.2	99	0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	120	0.00
8 S	Nitrobenzene-d5	0.419	0.368	12.2	117	0.00
9	Naphthalene	4.581	4.484	2.1	120	-0.01
10	Hexachlorobutadiene	0.303	0.279	7.9	114	0.00
11 SURR	2-Methylnaphthalene-d10	0.739	0.700	5.3	116	0.00
12	2-Methylnaphthalene	0.743	0.705	5.1	117	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	123	0.00
14 S	2,4,6-Tribromophenol	0.202	0.183	9.4	113	-0.01
15 S	2-Fluorobiphenyl	1.656	1.647	0.5	124	-0.01
16	Acenaphthylene	2.057	1.941	5.6	122	-0.01
17	Acenaphthene	1.220	1.176	3.6	120	-0.01
18	Fluorene	1.760	1.658	5.8	119	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	118	-0.01
20	4-Bromophenyl-phenylether	0.217	0.192	11.5	107	0.00
21	Hexachlorobenzene	0.258	0.241	6.6	117	0.00
22	Pentachlorophenol	0.052	0.050	3.8	155#	-0.01
23	Phenanthrene	1.137	1.037	8.8	110	-0.01
24	Anthracene	0.972	0.805	17.2	103	0.00
25 SURR	Fluoranthene-d10	6.385	5.588	12.5	107	0.00
26	Fluoranthene	1.605	1.397	13.0	107	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	109	-0.01
28	Pvrene	1.224	1.195	2.4	107	0.00
29 S	Terphenyl-d14	0.972	0.941	3.2	106	0.00
30	Benzo(a)anthracene	1.248	1.111	11.0	99	0.00
31	Chrysene	1.336	1.268	5.1	106	-0.01
32	Bis(2-ethylhexyl)phthalate	2.657	2.146	19.2	96	-0.01
33	Indeno(1,2,3-cd)pyrene	1.410	1.384	1.8	109	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	115	-0.01
35	Benzo(b)fluoranthene	1.315	1.235	6.1	111	0.00
36	Benzo(k)fluoranthene	1.377	1.291	6.2	113	-0.01
37 C	Benzo(a)pyrene	1.261	1.205	4.4	113	-0.01
38	Dibenzo(a,h)anthracene	1.193	1.081	9.4	105	-0.02
39	Benzo(a,h,i)perylene	1.248	1.155	7.5	108	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031319\
Data File : BE098923.D
Acq On : 13 Mar 2019 22:08
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 14 07:24:44 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 06 15:33:24 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			