

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031319\
 Data File : BE098936.D
 Acq On : 14 Mar 2019 7:26
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Mar 14 08:01:40 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 15:33:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	0.00
2	1,4-Dioxane	0.789	0.917	-16.2	141	0.00
3	n-Nitrosodimethylamine	1.006	0.607	39.7#	69	0.03
4 S	2-Fluorophenol	1.196	1.124	6.0	106	0.01
5 S	Phenol-d6	1.689	1.860	-10.1	129	0.00
6	bis(2-Chloroethyl)ether	1.304	1.091	16.3	101	0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	127	0.00
8 S	Nitrobenzene-d5	0.419	0.480	-14.6	162#	0.00
9	Naphthalene	4.581	4.444	3.0	127	-0.01
10	Hexachlorobutadiene	0.303	0.286	5.6	124	0.00
11 SURR	2-Methylnaphthalene-d10	0.739	0.666	9.9	117	0.00
12	2-Methylnaphthalene	0.743	0.683	8.1	121	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	131	0.00
14 S	2,4,6-Tribromophenol	0.202	0.207	-2.5	136	-0.01
15 S	2-Fluorobiphenyl	1.656	2.093	-26.4#	168#	-0.01
16	Acenaphthylene	2.057	1.953	5.1	130	-0.01
17	Acenaphthene	1.220	1.184	3.0	128	-0.01
18	Fluorene	1.760	1.682	4.4	128	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	122	-0.01
20	4-Bromophenyl-phenylether	0.217	0.219	-0.9	126	0.00
21	Hexachlorobenzene	0.258	0.257	0.4	128	0.00
22	Pentachlorophenol	0.052	0.047	9.6	150	-0.01
23	Phenanthrene	1.137	1.100	3.3	121	-0.01
24	Anthracene	0.972	0.898	7.6	119	0.00
25 SURR	Fluoranthene-d10	6.385	5.746	10.0	114	0.00
26	Fluoranthene	1.605	1.491	7.1	118	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	118	-0.01
28	Pvrene	1.224	1.243	-1.6	120	0.00
29 S	Terphenyl-d14	0.972	1.252	-28.8#	153#	0.00
30	Benzo(a)anthracene	1.248	1.125	9.9	108	0.00
31	Chrysene	1.336	1.300	2.7	117	-0.01
32	Bis(2-ethylhexyl)phthalate	2.657	2.126	20.0	103	-0.01
33	Indeno(1,2,3-cd)pyrene	1.410	1.360	3.5	116	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	122	-0.01
35	Benzo(b)fluoranthene	1.315	1.236	6.0	119	0.00
36	Benzo(k)fluoranthene	1.377	1.359	1.3	127	-0.02
37 C	Benzo(a)pyrene	1.261	1.212	3.9	122	-0.01
38	Dibenzo(a,h)anthracene	1.193	1.071	10.2	111	-0.02
39	Benzo(a,h,i)perylene	1.248	1.168	6.4	117	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031319\
Data File : BE098936.D
Acq On : 14 Mar 2019 7:26
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 14 08:01:40 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 06 15:33:24 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			