

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031319\
 Data File : BE098936.D
 Acq On : 14 Mar 2019 7:26
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 14 08:01:40 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 15:33:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	117	0.00
2	1,4-Dioxane	0.400	0.491	-22.7	141	0.00
3	n-Nitrosodimethylamine	0.400	0.241	39.8#	69	0.03
4 S	2-Fluorophenol	0.400	0.376	6.0	106	0.01
5 S	Phenol-d6	0.400	0.440	-10.0	129	0.00
6	bis(2-Chloroethyl)ether	0.400	0.335	16.3	101	0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	127	0.00
8 S	Nitrobenzene-d5	0.400	0.457	-14.2	162	0.00
9	Naphthalene	0.400	0.388	3.0	127	-0.01
10	Hexachlorobutadiene	0.400	0.377	5.8	124	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.360	10.0	117	0.00
12	2-Methylnaphthalene	0.400	0.367	8.3	121	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	131	0.00
14 S	2,4,6-Tribromophenol	0.400	0.410	-2.5	136	-0.01
15 S	2-Fluorobiphenyl	0.400	0.505	-26.2#	168	-0.01
16	Acenaphthylene	0.400	0.380	5.0	130	-0.01
17	Acenaphthene	0.400	0.388	3.0	128	-0.01
18	Fluorene	0.400	0.382	4.5	128	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	122	-0.01
20	4-Bromophenyl-phenylether	0.400	0.405	-1.3	126	0.00
21	Hexachlorobenzene	0.400	0.399	0.3	128	0.00
22	Pentachlorophenol	0.400	0.479	-19.7	150	-0.01
23	Phenanthrene	0.400	0.387	3.3	121	-0.01
24	Anthracene	0.400	0.370	7.5	119	0.00
25 SURR	Fluoranthene-d10	0.400	0.360	10.0	114	0.00
26	Fluoranthene	0.400	0.372	7.0	118	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	118	-0.01
28	Pvrene	0.400	0.406	-1.5	120	0.00
29 S	Terphenyl-d14	0.400	0.515	-28.7#	153	0.00
30	Benzo(a)anthracene	0.400	0.360	10.0	108	0.00
31	Chrysene	0.400	0.389	2.8	117	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.320	20.0	103	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.386	3.5	116	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	122	-0.01
35	Benzo(b)fluoranthene	0.400	0.376	6.0	119	0.00
36	Benzo(k)fluoranthene	0.400	0.395	1.3	127	-0.02
37 C	Benzo(a)pyrene	0.400	0.384	4.0	122	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.359	10.3	111	-0.02
39	Benzo(a,h,i)perylene	0.400	0.375	6.3	117	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031319\
Data File : BE098936.D
Acq On : 14 Mar 2019 7:26
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 14 08:01:40 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 06 15:33:24 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				