

Data Path : Z:\HPCHEM1\BNA E\DATA\BE031418\
 Data File : BE095792.D
 Acq On : 15 Mar 2018 7:58
 Operator : SJ/JU
 Sample : J1835-14
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :

Quant Time: Mar 15 09:16:27 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 15 05:35:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.44	152	2344	0.40	ng/ul	0.02
2) Naphthalene-d8	11.27	136	7143	0.40	ng/ul	0.02
6) Acenaphthene-d10	15.02	164	3992	0.40	ng/ul	0.02
10) Phenanthrene-d10	17.73	188	7828m	0.40	ng/ul	0.02
16) Chrysene-d12	21.83	240	7904m	0.40	ng/ul	0.03
20) Perylene-d12	24.50	264	8820m	0.40	ng/ul	0.07
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.82	152	3022	0.25	ng/ul	0.02
14) Fluoranthene-d10	19.71	212	7451m	0.30	ng/ul	0.02
Target Compounds				Ovalue		
3) Naphthalene	11.32	128	19818	0.994	ng/ul#	89
5) 2-Methylnaphthalene	12.89	142	36252	2.615	ng/ul	100
8) Acenaphthene	15.09	153	1545	0.102	ng/ul#	69
12) Phenanthrene	17.77	178	97361m	4.153	ng/ul	
13) Anthracene	17.87	178	10075m	0.460	ng/ul	
15) Fluoranthene	19.74	202	19912m	0.701	ng/ul	
17) Pyrene	20.09	202	94074	3.134	ng/ul#	77
18) Benzo(a)anthracene	21.81	228	31998m	1.255	ng/ul	
19) Chrysene	21.86	228	59340m	2.285	ng/ul	
21) Benzo(b)fluoranthene	23.67	252	15284m	0.500	ng/ul	
23) Benzo(a)pyrene	24.37	252	24208m	0.868	ng/ul	
24) Indeno(1,2,3-cd)pyrene	27.36	276	10724	0.347	ng/ul#	70
25) Dibenzo(a,h)anthracene	27.36	278	6459m	0.261	ng/ul	
26) Benzo(g,h,i)perylene	28.26	276	34525	1.299	ng/ul#	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE031418\
Data File : BE095792.D
Acq On : 15 Mar 2018 7:58
Operator : SJ/JU
Sample : J1835-14
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :

Quant Time: Mar 15 09:16:27 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 15 05:35:28 2018
Response via : Initial Calibration

