

Data Path : Z:\HPCHEM1\BNA_E\Data\BE031418\
 Data File : BE095787.D
 Acq On : 15 Mar 2018 4:56
 Operator : SJ/JU
 Sample : J1835-01
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F6F00

Manual Integrations
 APPROVED

Sohil
 3/15/2018 6:55:06 PM

Quant Time: Mar 16 02:44:54 2018
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE031318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 15 05:35:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.43	152	2302	0.40	ng/ul	0.00
2) Naphthalene-d8	11.25	136	7136	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.01	164	4346	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.72	188	8864m	0.40	ng/ul	0.00
16) Chrysene-d12	21.81	240	8325m	0.40	ng/ul	0.00
20) Perylene-d12	24.44	264	8288m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.81	152	2948	0.24	ng/ul	0.00
14) Fluoranthene-d10	19.69	212	6867m	0.24	ng/ul	0.00
Target Compounds				Qvalue		
3) Naphthalene	11.31	128	4645	0.233	ng/ul#	90
5) 2-Methylnaphthalene	12.88	142	4770	0.344	ng/ul	100
8) Acenaphthene	15.07	153	8156	0.494	ng/ul	94
9) Fluorene	16.04	166	7004	0.391	ng/ul	89
12) Phenanthrene	17.76	178	117552m	4.428	ng/ul	
13) Anthracene	17.85	178	26243m	1.057	ng/ul	
15) Fluoranthene	19.72	202	165024m	5.132	ng/ul	
17) Pyrene	20.08	202	144902m	4.584	ng/ul	
18) Benzo(a)anthracene	21.79	228	73797m	2.747	ng/ul	
19) Chrysene	21.84	228	71964	2.631	ng/ul	95
21) Benzo(b)fluoranthene	23.62	252	82537m	2.875	ng/ul	
22) Benzo(k)fluoranthene	23.67	252	30946m	1.134	ng/ul	
23) Benzo(a)pyrene	24.32	252	58316m	2.225	ng/ul	
24) Indeno(1,2,3-cd)pyrene	27.25	276	38839m	1.336	ng/ul	
25) Dibenzo(a,h)anthracene	27.26	278	10351m	0.445	ng/ul	
26) Benzo(g,h,i)perylene	28.14	276	45462	1.821	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\Data\BE031418\
Data File : BE095787.D
Acq On : 15 Mar 2018 4:56
Operator : SJ/JU
Sample : J1835-01
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F6F00

Manual Integrations
APPROVED

Sohil
3/15/2018 6:55:06 PM

Quant Time: Mar 16 02:44:54 2018
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE031318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 15 05:35:28 2018
Response via : Initial Calibration

