

Data Path : Z:\HPCHEM1\BNA E\Data\BE031418\
 Data File : BE095783.D
 Acq On : 15 Mar 2018 2:33
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.418

Quant Time: Mar 15 05:20:21 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 15 03:09:28 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	105	0.00
2 I	Naphthalene-d8	0.400	0.400	0.0	111	0.00
3	Naphthalene	0.400	0.366	8.5	100	0.00
4 SURR	2-Methylnaphthalene-d10	0.400	0.374	6.5	103	0.00
5	2-Methylnaphthalene	0.400	0.373	6.8	104	0.00
6 I	Acenaphthene-d10	0.400	0.400	0.0	119	0.00
7	Acenaphthylene	0.400	0.371	7.3	111	0.00
8 C	Acenaphthene	0.400	0.350	12.5	105	0.00
9	Fluorene	0.400	0.359	10.3	107	0.00
10 I	Phenanthrene-d10	0.400	0.400	0.0	117	0.00
11	Pentachlorophenol	0.400	0.543	-35.8#	204	0.00
12	Phenanthrene	0.400	0.351	12.3	102	0.00
13	Anthracene	0.400	0.379	5.3	110	0.00
14 SURR	Fluoranthene-d10	0.400	0.379	5.3	111	0.00
15 C	Fluoranthene	0.400	0.363	9.3	105	0.00
16 I	Chrysene-d12	0.400	0.400	0.0	117	0.00
17	Pyrene	0.400	0.411	-2.7	120	0.00
18	Benzo(a)anthracene	0.400	0.378	5.5	110	0.00
19	Chrysene	0.400	0.357	10.8	103	0.00
20 I	Perylene-d12	0.400	0.400	0.0	127	0.00
21	Benzo(b)fluoranthene	0.400	0.334	16.5	106	0.00
22	Benzo(k)fluoranthene	0.400	0.350	12.5	111	0.00
23 C	Benzo(a)pyrene	0.400	0.350	12.5	112	0.00
24	Indeno(1,2,3-cd)pyrene	0.400	0.363	9.3	116	0.00
25	Dibenzo(a,h)anthracene	0.400	0.370	7.5	118	0.00
26	Benzo(a,h,i)perylene	0.400	0.358	10.5	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0