

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031419\
 Data File : BE098938.D
 Acq On : 14 Mar 2019 10:12
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 15 05:33:16 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 15:33:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	1,4-Dioxane	0.789	0.759	3.8	94	0.00
3	n-Nitrosodimethylamine	1.006	0.813	19.2	75	0.04
4 S	2-Fluorophenol	1.196	1.078	9.9	82	0.02
5 S	Phenol-d6	1.689	1.471	12.9	82	0.02
6	bis(2-Chloroethyl)ether	1.304	1.051	19.4	79	0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	107	0.00
8 S	Nitrobenzene-d5	0.419	0.347	17.2	99	0.01
9	Naphthalene	4.581	4.449	2.9	107	0.00
10	Hexachlorobutadiene	0.303	0.296	2.3	108	0.00
11 SURR	2-Methylnaphthalene-d10	0.739	0.720	2.6	106	0.00
12	2-Methylnaphthalene	0.743	0.672	9.6	100	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	113	0.00
14 S	2,4,6-Tribromophenol	0.202	0.152	24.8	86	0.00
15 S	2-Fluorobiphenyl	1.656	1.532	7.5	106	0.00
16	Acenaphthylene	2.057	1.898	7.7	110	0.00
17	Acenaphthene	1.220	1.186	2.8	111	-0.01
18	Fluorene	1.760	1.637	7.0	108	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
20	4-Bromophenyl-phenylether	0.217	0.185	14.7	89	0.00
21	Hexachlorobenzene	0.258	0.250	3.1	105	0.00
22	Pentachlorophenol	0.052	0.025	51.9#	66	0.03
23	Phenanthrene	1.137	1.093	3.9	101	0.00
24	Anthracene	0.972	0.808	16.9	90	0.00
25 SURR	Fluoranthene-d10	6.385	5.591	12.4	93	0.00
26	Fluoranthene	1.605	1.401	12.7	93	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	90	0.00
28	Pvrene	1.224	1.248	-2.0	92	0.00
29 S	Terphenyl-d14	0.972	0.913	6.1	85	0.00
30	Benzo(a)anthracene	1.248	1.139	8.7	84	0.00
31	Chrysene	1.336	1.259	5.8	87	0.00
32	Bis(2-ethylhexyl)phthalate	2.657	2.050	22.8	76	-0.01
33	Indeno(1,2,3-cd)pyrene	1.410	1.252	11.2	81	0.00
34 I	Perylene-d12	1.000	1.000	0.0	92	0.00
35	Benzo(b)fluoranthene	1.315	1.188	9.7	86	0.00
36	Benzo(k)fluoranthene	1.377	1.382	-0.4	97	0.00
37 C	Benzo(a)pyrene	1.261	1.186	5.9	89	0.00
38	Dibenzo(a,h)anthracene	1.193	0.977	18.1	76	0.00
39	Benzo(a,h,i)perylene	1.248	1.079	13.5	81	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031419\
Data File : BE098938.D
Acq On : 14 Mar 2019 10:12
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 15 05:33:16 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 06 15:33:24 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			