

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031419\
 Data File : BE098965.D
 Acq On : 15 Mar 2019 4:31
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 15 06:06:51 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 15 05:31:33 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	120	0.00
2	1,4-Dioxane	0.400	0.353	11.8	111	0.00
3	n-Nitrosodimethylamine	0.400	0.311	22.3	92	0.01
4 S	2-Fluorophenol	0.400	0.366	8.5	106	0.01
5 S	Phenol-d6	0.400	0.387	3.3	116	0.00
6	bis(2-Chloroethyl)ether	0.400	0.355	11.3	111	0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	130	0.00
8 S	Nitrobenzene-d5	0.400	0.319	20.3	115	0.00
9	Naphthalene	0.400	0.392	2.0	130	-0.01
10	Hexachlorobutadiene	0.400	0.378	5.5	127	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.386	3.5	128	0.00
12	2-Methylnaphthalene	0.400	0.389	2.8	130	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	135	0.00
14 S	2,4,6-Tribromophenol	0.400	0.313	21.8	107	0.00
15 S	2-Fluorobiphenyl	0.400	0.379	5.3	129	-0.01
16	Acenaphthylene	0.400	0.383	4.3	136	-0.01
17	Acenaphthene	0.400	0.391	2.3	133	-0.01
18	Fluorene	0.400	0.369	7.8	128	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	120	0.00
20	4-Bromophenyl-phenylether	0.400	0.383	4.3	117	0.00
21	Hexachlorobenzene	0.400	0.383	4.3	121	0.00
22	Pentachlorophenol	0.400	0.450	-12.5	127	0.00
23	Phenanthrene	0.400	0.372	7.0	114	0.00
24	Anthracene	0.400	0.343	14.2	109	0.00
25 SURR	Fluoranthene-d10	0.400	0.337	15.8	105	0.00
26	Fluoranthene	0.400	0.338	15.5	105	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	102	-0.01
28	Pvrene	0.400	0.407	-1.7	104	0.00
29 S	Terphenyl-d14	0.400	0.381	4.8	98	0.00
30	Benzo(a)anthracene	0.400	0.349	12.8	91	0.00
31	Chrysene	0.400	0.400	0.0	105	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.347	13.3	97	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.375	6.3	97	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	106	-0.01
35	Benzo(b)fluoranthene	0.400	0.382	4.5	105	0.00
36	Benzo(k)fluoranthene	0.400	0.381	4.8	106	-0.01
37 C	Benzo(a)pyrene	0.400	0.392	2.0	108	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.312	22.0	84	-0.01
39	Benzo(a,h,i)perylene	0.400	0.367	8.3	100	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031419\
Data File : BE098965.D
Acq On : 15 Mar 2019 4:31
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 15 06:06:51 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 15 05:31:33 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				