

Data Path : Z:\HPCHEM1\BNA E\DATA\BE031518\
 Data File : BE095807.D
 Acq On : 15 Mar 2018 17:16
 Operator : SJ/JU
 Sample : J1835-01DL 10X
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F6F00DL

Manual Integrations
 APPROVED

Sohil
 3/16/2018 12:03:44 PM

Quant Time: Mar 16 03:00:07 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 16 02:40:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.45	152	1975	0.40	ng/ul	0.00
2) Naphthalene-d8	11.28	136	5637	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.04	164	3026	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.74	188	6338m	0.40	ng/ul	0.00
16) Chrysene-d12	21.84	240	6297m	0.40	ng/ul	0.00
20) Perylene-d12	24.49	264	6434m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.83	152	238m	0.02	ng/ul	0.00
14) Fluoranthene-d10	19.72	212	529m	0.03	ng/ul	0.00
Target Compounds						
					Ovalue	
3) Naphthalene	11.32	128	381	0.024	ng/ul#	56
5) 2-Methylnaphthalene	12.90	142	371	0.034	ng/ul	100
8) Acenaphthene	15.10	153	576	0.050	ng/ul#	91
9) Fluorene	16.07	166	511	0.041	ng/ul#	80
12) Phenanthrene	17.78	178	8517m	0.449	ng/ul	
13) Anthracene	17.87	178	1861m	0.105	ng/ul	
15) Fluoranthene	19.75	202	12162m	0.529	ng/ul	
17) Pyrene	20.10	202	10908	0.456	ng/ul#	77
18) Benzo(a)anthracene	21.81	228	5576m	0.274	ng/ul	
19) Chrysene	21.87	228	5868m	0.284	ng/ul	
21) Benzo(b)fluoranthene	23.67	252	5769m	0.259	ng/ul	
22) Benzo(k)fluoranthene	23.72	252	2628m	0.124	ng/ul	
23) Benzo(a)pyrene	24.37	252	4624	0.227	ng/ul#	34
24) Indeno(1,2,3-cd)pyrene	27.35	276	2363	0.105	ng/ul#	75
25) Dibenzo(a,h)anthracene	27.34	278	756	0.042	ng/ul#	1
26) Benzo(a,h,i)perylene	28.25	276	2760	0.142	ng/ul#	36

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE031518\
Data File : BE095807.D
Acq On : 15 Mar 2018 17:16
Operator : SJ/JU
Sample : J1835-01DL 10X
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F6F00DL

Manual Integrations
APPROVED

Sohil
3/16/2018 12:03:44 PM

Quant Time: Mar 16 03:00:07 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 16 02:40:12 2018
Response via : Initial Calibration

