

Data Path : Z:\HPCHEM1\BNA E\Data\BE031518\
 Data File : BE095806.D
 Acq On : 15 Mar 2018 16:40
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.421

Quant Time: Mar 16 01:14:21 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 16 01:12:14 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	121	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	118	0.00
3	Naphthalene	1.117	1.010	9.6	105	0.00
4 SURR	2-Methylnaphthalene-d10	0.677	0.601	11.2	104	0.00
5	2-Methylnaphthalene	0.776	0.686	11.6	105	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	112	0.00
7	Acenaphthylene	2.020	1.948	3.6	108	0.00
8 C	Acenaphthene	1.521	1.360	10.6	101	0.00
9	Fluorene	1.651	1.457	11.8	99	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	106	0.00
11	Pentachlorophenol	0.056	0.052	7.1	124	0.00
12	Phenanthrene	1.198	1.059	11.6	93	0.00
13	Anthracene	1.120	1.061	5.3	99	0.00
14 SURR	Fluoranthene-d10	1.271	1.186	6.7	99	0.00
15 C	Fluoranthene	1.451	1.307	9.9	94	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
17	Pyrene	1.519	1.287	15.3	91	0.00
18	Benzo(a)anthracene	1.291	1.232	4.6	102	0.00
19	Chrysene	1.314	1.147	12.7	93	0.00
20 I	Perylene-d12	1.000	1.000	0.0	119	0.00
21	Benzo(b)fluoranthene	1.386	1.224	11.7	105	0.00
22	Benzo(k)fluoranthene	1.317	1.104	16.2	99	0.00
23 C	Benzo(a)pyrene	1.265	1.123	11.2	106	0.00
24	Indeno(1,2,3-cd)pyrene	1.403	1.190	15.2	101	0.00
25	Dibenzo(a,h)anthracene	1.122	1.025	8.6	109	0.00
26	Benzo(a,h,i)perylene	1.205	0.896	25.6#	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0