

Data Path : Z:\HPCHEM1\BNA E\Data\BE031518\
 Data File : BE095813.D
 Acq On : 15 Mar 2018 20:51
 Operator : SJ/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.422

Quant Time: Mar 16 04:03:11 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 16 02:40:12 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	128	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	127	0.00
3	Naphthalene	1.117	1.013	9.3	114	0.00
4 SURR	2-Methylnaphthalene-d10	0.677	0.596	12.0	111	0.00
5	2-Methylnaphthalene	0.776	0.685	11.7	112	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	123	0.00
7	Acenaphthylene	2.020	1.907	5.6	117	0.00
8 C	Acenaphthene	1.521	1.326	12.8	108	0.00
9	Fluorene	1.651	1.445	12.5	108	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	113	0.00
11	Pentachlorophenol	0.056	0.048	14.3	124	0.00
12	Phenanthrene	1.198	1.060	11.5	99	0.00
13	Anthracene	1.120	1.071	4.4	108	0.00
14 SURR	Fluoranthene-d10	1.271	1.212	4.6	108	0.00
15 C	Fluoranthene	1.451	1.322	8.9	102	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	116	0.00
17	Pyrene	1.519	1.302	14.3	99	0.00
18	Benzo(a)anthracene	1.291	1.227	5.0	109	0.00
19	Chrysene	1.314	1.147	12.7	100	0.00
20 I	Perylene-d12	1.000	1.000	0.0	126	0.00
21	Benzo(b)fluoranthene	1.386	1.217	12.2	111	0.00
22	Benzo(k)fluoranthene	1.317	1.146	13.0	109	0.00
23 C	Benzo(a)pyrene	1.265	1.142	9.7	114	-0.01
24	Indeno(1,2,3-cd)pyrene	1.403	1.192	15.0	108	-0.01
25	Dibenzo(a,h)anthracene	1.122	0.981	12.6	110	0.00
26	Benzo(a,h,i)perylene	1.205	0.886	26.5#	92	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0