

Data Path : Z:\HPCHEM1\BNA E\Data\BE031518\
 Data File : BE095813.D
 Acq On : 15 Mar 2018 20:51
 Operator : SJ/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.422

Quant Time: Mar 16 04:03:11 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 16 02:40:12 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	128	0.00
2 I	Naphthalene-d8	0.400	0.400	0.0	127	0.00
3	Naphthalene	0.400	0.363	9.3	114	0.00
4 SURR	2-Methylnaphthalene-d10	0.400	0.352	12.0	111	0.00
5	2-Methylnaphthalene	0.400	0.353	11.8	112	0.00
6 I	Acenaphthene-d10	0.400	0.400	0.0	123	0.00
7	Acenaphthylene	0.400	0.378	5.5	117	0.00
8 C	Acenaphthene	0.400	0.349	12.8	108	0.00
9	Fluorene	0.400	0.350	12.5	108	0.00
10 I	Phenanthrene-d10	0.400	0.400	0.0	113	0.00
11	Pentachlorophenol	0.400	0.341	14.8	124	0.00
12	Phenanthrene	0.400	0.354	11.5	99	0.00
13	Anthracene	0.400	0.382	4.5	108	0.00
14 SURR	Fluoranthene-d10	0.400	0.382	4.5	108	0.00
15 C	Fluoranthene	0.400	0.364	9.0	102	0.00
16 I	Chrysene-d12	0.400	0.400	0.0	116	0.00
17	Pyrene	0.400	0.343	14.2	99	0.00
18	Benzo(a)anthracene	0.400	0.380	5.0	109	0.00
19	Chrysene	0.400	0.349	12.8	100	0.00
20 I	Perylene-d12	0.400	0.400	0.0	126	0.00
21	Benzo(b)fluoranthene	0.400	0.351	12.3	111	0.00
22	Benzo(k)fluoranthene	0.400	0.348	13.0	109	0.00
23 C	Benzo(a)pyrene	0.400	0.361	9.8	114	-0.01
24	Indeno(1,2,3-cd)pyrene	0.400	0.340	15.0	108	-0.01
25	Dibenzo(a,h)anthracene	0.400	0.350	12.5	110	0.00
26	Benzo(a,h,i)perylene	0.400	0.294	26.5#	92	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0