

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE031616\
 Data File : BE091456.D
 Acq On : 16 Mar 2016 23:06
 Operator : UM/SJ
 Sample : SSTDICC005
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC005

Manual Integrations
 APPROVED

Sohil
 3/17/2016 5:22:23 PM

Quant Time: Mar 17 15:38:33 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE031616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 07:07:36 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	65973	5.00	ng	0.00
4) Naphthalene-d8	10.61	136	307390	5.00	ng	0.00
8) Acenaphthene-d10	14.44	164	180219	5.00	ng	0.00
13) Phenanthrene-d10	17.18	188	430963	5.00	ng	0.00
19) Chrysene-d12	21.34	240	556898	5.00	ng	0.00
25) Perylene-d12	23.81	264	528821	5.00	ng	0.00
System Monitoring Compounds						
5) Nitrobenzene-d5	8.99	82	81207	5.67	ng	0.00
9) 2-Fluorobiphenyl	13.08	172	252651	5.78	ng	0.00
21) Terphenyl-d14	19.80	244	434259	5.17	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.37	88	39244	6.45	ng	# 100
3) n-Nitrosodimethylamine	3.66	42	72554	10.31	ng	# 100
6) Naphthalene	10.67	128	394778	5.57	ng	99
7) 2-Methylnaphthalene	12.27	142	264429	6.27	ng	93
10) Acenaphthylene	14.16	152	468306	1.75	ng	# 100
11) Acenaphthene	14.51	154	298189	5.49	ng	# 100
12) Fluorene	15.49	166	381053	5.42	ng	# 100
14) Hexachlorobenzene	16.48	284	102607	4.19	ng	# 100
15) Pentachlorophenol	16.83	266	52166	9.68	ng	# 83
16) Phenanthrene	17.23	178	646697	4.40	ng	# 100
17) Anthracene	17.31	178	622233	2.96	ng	# 100
18) Fluoranthene	19.22	202	775883	3.63	ng	# 68
20) Pyrene	19.59	202	849784	5.40	ng	# 92
22) Benzo(a)anthracene	21.31	228	982208	5.10	ng	99
24) Indeno(1,2,3-cd)pyrene	26.40	276	971579	5.70	ng	# 100
26) Benzo(b)fluoranthene	23.04	252	856744	5.55	ng	92
27) Benzo(k)fluoranthene	23.10	252	913268	6.40	ng	# 93
28) Benzo(a)pyrene	23.69	252	831159	6.38	ng	# 92
29) Dibenzo(a,h)anthracene	26.43	278	817755	6.54	ng	96
30) Benzo(g,h,i)perylene	27.20	276	820856	6.13	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE031616\
Data File : BE091456.D
Acq On : 16 Mar 2016 23:06
Operator : UM/SJ
Sample : SSTDICC005
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Sohil
3/17/2016 5:22:23 PM

Quant Time: Mar 17 15:38:33 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE031616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 17 07:07:36 2016
Response via : Initial Calibration

