

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE031616\
 Data File : BE091458.D
 Acq On : 17 Mar 2016 00:21
 Operator : UM/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE031616

Quant Time: Mar 17 15:39:23 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE031616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 10:23:05 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
2	1,4-Dioxane	0.646	0.791	-22.4	105	0.00
3	n-Nitrosodimethylamine	0.960	1.115	-16.1	106	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	83	0.00
5 S	Nitrobenzene-d5	0.227	0.271	-19.4	116	0.00
6	Naphthalene	1.228	1.520	-23.8	107	0.00
7	2-Methylnaphthalene	0.801	0.961	-20.0	107	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	81	0.00
9 S	2-Fluorobiphenyl	1.291	1.569	-21.5	107	0.00
10	Acenaphthylene	2.178	2.387	-9.6	108	0.00
11	Acenaphthene	1.537	1.905	-23.9	110	0.00
12	Fluorene	1.985	2.411	-21.5	106	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	81	0.00
14	Hexachlorobenzene	0.219	0.264	-20.5	106	-0.02
15	Pentachlorophenol	0.090	0.097	-7.8	106	-0.02
16	Phenanthrene	1.430	1.760	-23.1	106	0.00
17	Anthracene	1.275	1.486	-16.5	105	0.00
18	Fluoranthene	1.631	1.966	-20.5	105	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	90	0.00
20	Pyrene	1.427	1.542	-8.1	106	0.00
21 S	Terphenyl-d14	0.743	0.783	-5.4	99	0.00
22	Benzo(a)anthracene	1.752	1.571	10.3	88	0.00
23	Chrysene	1.525	1.826	-19.7	109	0.00
24	Indeno(1,2,3-cd)pyrene	1.673	1.797	-7.4	104	0.00
25 I	Perylene-d12	1.000	1.000	0.0	83	0.00
26	Benzo(b)fluoranthene	1.703	2.030	-19.2	104	-0.01
27	Benzo(k)fluoranthene	1.579	1.751	-10.9	101	0.00
28 C	Benzo(a)pyrene	1.449	1.629	-12.4	103	0.00
29	Dibenzo(a,h)anthracene	1.406	1.609	-14.4	105	0.00
30	Benzo(g,h,i)perylene	1.471	1.722	-17.1	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0