

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE031616\
 Data File : BE091458.D
 Acq On : 17 Mar 2016 00:21
 Operator : UM/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE031616

Quant Time: Mar 17 15:39:23 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE031616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 10:23:05 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	83	0.00
2	1,4-Dioxane	0.400	0.490	-22.5	105	0.00
3	n-Nitrosodimethylamine	0.400	0.465	-16.3	106	0.00
4 I	Naphthalene-d8	5.000	5.000	0.0	83	0.00
5 S	Nitrobenzene-d5	0.400	0.478	-19.5	116	0.00
6	Naphthalene	0.400	0.495	-23.7	107	0.00
7	2-Methylnaphthalene	0.400	0.480	-20.0	107	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	81	0.00
9 S	2-Fluorobiphenyl	0.400	0.486	-21.5	107	0.00
10	Acenaphthylene	0.400	0.438	-9.5	108	0.00
11	Acenaphthene	0.400	0.496	-24.0	110	0.00
12	Fluorene	0.400	0.486	-21.5	106	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	81	0.00
14	Hexachlorobenzene	0.400	0.483	-20.7	106	-0.02
15	Pentachlorophenol	0.400	0.562	-40.5#	106	-0.02
16	Phenanthrene	0.400	0.492	-23.0	106	0.00
17	Anthracene	0.400	0.466	-16.5	105	0.00
18	Fluoranthene	0.400	0.482	-20.5	105	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	90	0.00
20	Pyrene	0.400	0.432	-8.0	106	0.00
21 S	Terphenyl-d14	0.400	0.422	-5.5	99	0.00
22	Benzo(a)anthracene	0.400	0.359	10.3	88	0.00
23	Chrysene	0.400	0.533	-33.3#	109	0.00
24	Indeno(1,2,3-cd)pyrene	0.400	0.430	-7.5	104	0.00
25 I	Perylene-d12	5.000	5.000	0.0	83	0.00
26	Benzo(b)fluoranthene	0.400	0.477	-19.2	104	-0.01
27	Benzo(k)fluoranthene	0.400	0.443	-10.7	101	0.00
28 C	Benzo(a)pyrene	0.400	0.449	-12.2	103	0.00
29	Dibenzo(a,h)anthracene	0.400	0.458	-14.5	105	0.00
30	Benzo(g,h,i)perylene	0.400	0.468	-17.0	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0