

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE031716\
 Data File : BE091465.D
 Acq On : 17 Mar 2016 14:10
 Operator : UM/SJ
 Sample : H1746-01MS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 TILCON-MH-0.75STONE-001MS

Quant Time: Mar 17 16:17:08 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE031616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 15:28:00 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	61926	5.00	ng	0.00
4) Naphthalene-d8	10.61	136	292027	5.00	ng	0.00
8) Acenaphthene-d10	14.45	164	180764	5.00	ng	0.00
13) Phenanthrene-d10	17.18	188	425663	5.00	ng	0.00
19) Chrysene-d12	21.33	240	545585	5.00	ng	0.00
25) Perylene-d12	23.81	264	523895	5.00	ng	0.00
System Monitoring Compounds						
5) Nitrobenzene-d5	8.98	82	65839	4.16	ng	0.00
9) 2-Fluorobiphenyl	13.07	172	229435	4.92	ng	-0.01
21) Terphenyl-d14	19.80	244	349602	4.31	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.37	88	26971	3.72	ng	# 100
3) n-Nitrosodimethylamine	3.66	42	38200	3.21	ng	# 100
6) Naphthalene	10.67	128	254716	3.55	ng	98
7) 2-Methylnaphthalene	12.27	142	186224	3.98	ng	91
10) Acenaphthylene	14.16	152	294862	3.74	ng	# 100
11) Acenaphthene	14.50	154	199648	3.59	ng	# 100
12) Fluorene	15.49	166	258616	3.60	ng	# 100
14) Hexachlorobenzene	16.48	284	71073	3.81	ng	# 100
15) Pentachlorophenol	16.83	266	55506	5.53	ng	90
16) Phenanthrene	17.22	178	488857	4.02	ng	# 100
17) Anthracene	17.31	178	425890	3.92	ng	# 100
18) Fluoranthene	19.23	202	665242	4.79	ng	# 69
20) Pyrene	19.59	202	670191	4.31	ng	# 92
22) Benzo(a)anthracene	21.31	228	776476	4.06	ng	97
23) Chrysene	21.38	228	408117m	3.25	ng	
24) Indeno(1,2,3-cd)pyrene	26.40	276	670393	3.67	ng	# 100
26) Benzo(b)fluoranthene	23.09	252	626775	3.51	ng	92
27) Benzo(k)fluoranthene	23.09	252	626775	3.79	ng	# 92
28) Benzo(a)pyrene	23.69	252	608123	4.00	ng	# 92
29) Dibenzo(a,h)anthracene	26.42	278	546184	3.71	ng	96
30) Benzo(g,h,i)perylene	27.18	276	574318	3.73	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE031716\
Data File : BE091465.D
Acq On : 17 Mar 2016 14:10
Operator : UM/SJ
Sample : H1746-01MS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
TILCON-MH-0.75STONE-001MS

Quant Time: Mar 17 16:17:08 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE031616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 17 15:28:00 2016
Response via : Initial Calibration

