

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE031716\
 Data File : BE091467.D
 Acq On : 17 Mar 2016 15:25
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 17 16:19:00 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE031616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 15:28:00 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2	1,4-Dioxane	0.646	0.593	8.2	82	0.00
3	n-Nitrosodimethylamine	0.960	0.831	13.4	82	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	87	0.00
5 S	Nitrobenzene-d5	0.227	0.216	4.8	98	-0.01
6	Naphthalene	1.228	1.152	6.2	86	-0.02
7	2-Methylnaphthalene	0.801	0.733	8.5	86	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	85	0.00
9 S	2-Fluorobiphenyl	1.291	1.183	8.4	86	0.00
10	Acenaphthylene	2.178	1.839	15.6	88	0.00
11	Acenaphthene	1.537	1.506	2.0	92	0.00
12	Fluorene	1.985	1.859	6.3	86	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	86	0.00
14	Hexachlorobenzene	0.219	0.202	7.8	86	-0.02
15	Pentachlorophenol	0.090	0.079	12.2	91	-0.02
16	Phenanthrene	1.430	1.362	4.8	87	-0.01
17	Anthracene	1.275	1.155	9.4	86	-0.01
18	Fluoranthene	1.631	1.574	3.5	89	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
20	Pyrene	1.427	1.209	15.3	89	0.00
21 S	Terphenyl-d14	0.743	0.607	18.3	83	0.00
22	Benzo(a)anthracene	1.752	1.369	21.9	82	0.00
23	Chrysene	1.525	1.375	9.8	88	0.00
24	Indeno(1,2,3-cd)pyrene	1.673	1.421	15.1	88	0.00
25 I	Perylene-d12	1.000	1.000	0.0	89	0.00
26	Benzo(b)fluoranthene	1.703	1.471	13.6	81	-0.01
27	Benzo(k)fluoranthene	1.579	1.516	4.0	94	-0.01
28 C	Benzo(a)pyrene	1.449	1.300	10.3	88	0.00
29	Dibenzo(a,h)anthracene	1.406	1.276	9.2	89	0.00
30	Benzo(g,h,i)perylene	1.471	1.337	9.1	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0