

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE031716\
 Data File : Be091461.D
 Acq On : 17 Mar 2016 10:14
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 17 16:04:58 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE031616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 15:28:00 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2	1,4-Dioxane	0.646	0.591	8.5	88	0.00
3	n-Nitrosodimethylamine	0.960	0.807	15.9	86	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	93	0.00
5 S	Nitrobenzene-d5	0.227	0.191	15.9	92	0.00
6	Naphthalene	1.228	1.156	5.9	91	0.00
7	2-Methylnaphthalene	0.801	0.718	10.4	90	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	88	0.00
9 S	2-Fluorobiphenyl	1.291	1.195	7.4	89	0.00
10	Acenaphthylene	2.178	1.769	18.8	87	0.00
11	Acenaphthene	1.537	1.470	4.4	92	0.00
12	Fluorene	1.985	1.836	7.5	87	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	88	0.00
14	Hexachlorobenzene	0.219	0.199	9.1	87	-0.02
15	Pentachlorophenol	0.090	0.075	16.7	89	-0.02
16	Phenanthrene	1.430	1.350	5.6	89	0.00
17	Anthracene	1.275	1.108	13.1	85	-0.01
18	Fluoranthene	1.631	1.514	7.2	88	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
20	Pyrene	1.427	1.271	10.9	92	0.00
21 S	Terphenyl-d14	0.743	0.626	15.7	84	0.00
22	Benzo(a)anthracene	1.752	1.585	9.5	93	0.00
23	Chrysene	1.525	1.225	19.7	77	0.00
24	Indeno(1,2,3-cd)pyrene	1.673	1.474	11.9	90	0.00
25 I	Perylene-d12	1.000	1.000	0.0	92	0.00
26	Benzo(b)fluoranthene	1.703	1.594	6.4	91	-0.01
27	Benzo(k)fluoranthene	1.579	1.364	13.6	87	0.00
28 C	Benzo(a)pyrene	1.449	1.263	12.8	89	0.00
29	Dibenzo(a,h)anthracene	1.406	1.240	11.8	90	0.00
30	Benzo(g,h,i)perylene	1.471	1.332	9.4	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0