

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE031716\
 Data File : Be091461.D
 Acq On : 17 Mar 2016 10:14
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 17 16:04:58 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE031616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 15:28:00 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	93	0.00
2	1,4-Dioxane	0.400	0.364	9.0	88	0.00
3	n-Nitrosodimethylamine	0.400	0.336	16.0	86	0.00
4 I	Naphthalene-d8	5.000	5.000	0.0	93	0.00
5 S	Nitrobenzene-d5	0.400	0.479	-19.7	92	0.00
6	Naphthalene	0.400	0.377	5.8	91	0.00
7	2-Methylnaphthalene	0.400	0.359	10.3	90	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	88	0.00
9 S	2-Fluorobiphenyl	0.400	0.370	7.5	89	0.00
10	Acenaphthylene	0.400	0.325	18.8	87	0.00
11	Acenaphthene	0.400	0.383	4.3	92	0.00
12	Fluorene	0.400	0.370	7.5	87	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	88	0.00
14	Hexachlorobenzene	0.400	0.363	9.3	87	-0.02
15	Pentachlorophenol	0.400	0.491	-22.7	89	-0.02
16	Phenanthrene	0.400	0.378	5.5	89	0.00
17	Anthracene	0.400	0.348	13.0	85	-0.01
18	Fluoranthene	0.400	0.371	7.3	88	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	94	0.00
20	Pyrene	0.400	0.356	11.0	92	0.00
21 S	Terphenyl-d14	0.400	0.337	15.8	84	0.00
22	Benzo(a)anthracene	0.400	0.362	9.5	93	0.00
23	Chrysene	0.400	0.316	21.0	77	0.00
24	Indeno(1,2,3-cd)pyrene	0.400	0.353	11.8	90	0.00
25 I	Perylene-d12	5.000	5.000	0.0	92	0.00
26	Benzo(b)fluoranthene	0.400	0.374	6.5	91	-0.01
27	Benzo(k)fluoranthene	0.400	0.345	13.8	87	0.00
28 C	Benzo(a)pyrene	0.400	0.349	12.8	89	0.00
29	Dibenzo(a,h)anthracene	0.400	0.353	11.8	90	0.00
30	Benzo(g,h,i)perylene	0.400	0.362	9.5	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0