

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE031816\
 Data File : BE097483.D
 Acq On : 18 Mar 2016 1:57
 Operator : UM/SJ
 Sample : MDL-06-W
 Misc : 0.08PPM
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 MDL-03-W

Quant Time: Mar 18 04:57:10 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE031816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 18 04:36:48 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.83	152	43844	5.00	ng	0.00
6) Naphthalene-d8	10.61	136	197740	5.00	ng	0.00
10) Acenaphthene-d10	14.44	164	119137	5.00	ng	0.00
16) Phenanthrene-d10	17.18	188	292022	5.00	ng	0.00
22) Chrysene-d12	21.34	240	347908	5.00	ng	0.00
28) Perylene-d12	23.80	264	327753	5.00	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.42	112	84192	8.36	ng	0.00
5) Phenol-d6	6.99	99	103075	7.84	ng	0.00
7) Nitrobenzene-d5	8.98	82	37625	3.54	ng	0.00
11) 2,4,6-Tribromophenol	15.92	330	23951	5.53	ng	0.00
12) 2-Fluorobiphenyl	13.07	172	122575	4.09	ng	0.00
24) Terphenyl-d14	19.79	244	222516	5.32	ng	-0.01


Target Compounds						Qvalue
2) 1,4-Dioxane	3.37	88	956	0.17	ng	# 54
3) n-Nitrosodimethylamine	3.68	42	80	0.01	ng	# 1
8) Naphthalene	10.66	128	2897	0.06	ng	# 85
9) 2-Methylnaphthalene	12.27	142	1635	0.05	ng	# 83
13) Acenaphthylene	14.17	152	1709	0.16	ng	# 83
14) Acenaphthene	14.50	154	1773	0.05	ng	# 72
15) Fluorene	15.49	166	1985	0.04	ng	# 76
17) Hexachlorobenzene	16.48	284	730	0.06	ng	# 63
18) Pentachlorophenol	16.83	266	346	0.35	ng	# 31
19) Phenanthrene	17.22	178	5146	0.06	ng	# 84
20) Anthracene	17.32	178	3522	0.05	ng	# 68
21) Fluoranthene	19.23	202	4896	0.05	ng	# 85
23) Pyrene	19.59	202	5354	0.06	ng	95
25) Benzo(a)anthracene	21.33	228	6479	0.07	ng	96
26) Chrysene	21.38	228	6889	Below Cal	#	85
27) Indeno(1,2,3-cd)pyrene	26.38	276	2055	0.02	ng	# 68
29) Benzo(b)fluoranthene	23.05	252	6694	0.06	ng	# 85
30) Benzo(k)fluoranthene	23.09	252	5791	0.05	ng	# 93
31) Benzo(a)pyrene	23.69	252	4941	0.05	ng	# 69
32) Dibenzo(a,h)anthracene	26.42	278	3387	0.04	ng	# 63
33) Benzo(g,h,i)perylene	27.19	276	4221	0.05	ng	# 79


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE031816\
Data File : BE097483.D
Acq On : 18 Mar 2016 1:57
Operator : UM/SJ
Sample : MDL-06-W
Misc : 0.08PPM
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
MDL-03-W

Quant Time: Mar 18 04:57:10 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE031816.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 18 04:36:48 2016
Response via : Initial Calibration









#1

1,4-Dichlorobenzene-d4
Concen: 5.00 ng
RT: 7.83 min Scan# 892
Delta R.T. -0.01 min
Lab File: BE097483.D
Acq: 18 Mar 2016 1:57

Instrument :
BNA_E
ClientSampleId :
MDL-03-W




Tgt Ion:152 Resp: 43844
Ion Ratio Lower Upper
152 100
150 167.9 122.2 183.2
115 61.5 45.9 68.9




#2

1,4-Dioxane
Concen: 0.17 ng
RT: 3.37 min Scan# 133
Delta R.T. 0.00 min
Lab File: BE097483.D
Acq: 18 Mar 2016 1:57

Tgt Ion: 88 Resp: 956
Ion Ratio Lower Upper
88 100
58 41.1 61.0 91.6#
43 0.0 25.4 38.2#






#3




n-Nitrosodimethylamine
 Concen: 0.01 ng
 RT: 3.68 min Scan# 186
 Delta R.T. 0.00 min
 Lab File: BE097483.D
 Acq: 18 Mar 2016 1:57

Instrument :
 BNA_E
 ClientSampled :
 MDL-03-W

Tgt Ion: 42 Resp: 80
 Ion Ratio Lower Upper
 42 100
 74 0.0 84.3 126.5#
 44 551.3 6.7 10.1#




Abundance Ion 42.00 (41.70 to 42.70): BE0
 Ion 74.00 (73.70 to 74.70): BE0
 Ion 44.00 (43.70 to 44.70): BE0





#4


2-Fluorophenol
 Concen: 8.36 ng
 RT: 5.42 min Scan# 482
 Delta R.T. -0.01 min
 Lab File: BE097483.D
 Acq: 18 Mar 2016 1:57

Tgt Ion: 112 Resp: 84192
 Ion Ratio Lower Upper
 112 100
 64 60.2 53.3 79.9
 63 32.2 29.8 44.8



Abundance Ion 112.00 (111.70 to 112.70): E
 Ion 64.00 (63.70 to 64.70): BE0
 Ion 63.00 (62.70 to 63.70): BE0





#5

Phenol-d6

Concen: 7.84 ng

RT: 6.99 min Scan# 749

Delta R.T. -0.00 min

Lab File: BE097483.D

Acq: 18 Mar 2016 1:57

Instrument :

BNA_E

ClientSampleId :

MDL-03-W





Tgt Ion: 99 Resp: 103075

Ion Ratio Lower Upper

99 100

42 20.7 20.6 30.8

71 34.2 29.0 43.4



#6

Naphthalene-d8

Concen: 5.00 ng

RT: 10.61 min Scan# 1366

Delta R.T. 0.00 min

Lab File: BE097483.D

Acq: 18 Mar 2016 1:57

Tgt Ion: 136 Resp: 197740




Ion Ratio Lower Upper


136 100

137 9.9 8.8 13.2

54 8.2 5.2 7.8#

68 6.7 4.1 6.1#





#7

Nitrobenzene-d5

Concen: 3.54 ng

RT: 8.98 min Scan# 1088

Delta R.T. -0.00 min

Lab File: BE097483.D

Acq: 18 Mar 2016 1:57

Instrument :

BNA_E

ClientSampled :

MDL-03-W


Tgt Ion: 82 Resp: 37625

Ion Ratio Lower Upper

82 100

128 34.9 23.5 35.3

54 60.3 52.4 78.6




Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): BE0

Ion 54.00 (53.70 to 54.70): BE0

Time-->




Abundance

Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): BE0

Ion 54.00 (53.70 to 54.70): BE0

Time-->



#8

Naphthalene

Concen: 0.06 ng

RT: 10.66 min Scan# 1374

Delta R.T. 0.01 min

Lab File: BE097483.D

Acq: 18 Mar 2016 1:57


Tgt Ion: 128 Resp: 2897

Ion Ratio Lower Upper

128 100

129 7.5 8.2 12.2#

127 22.9 11.8 17.8#




Abundance Ion 128.00 (127.70 to 128.70): BE0

Ion 129.00 (128.70 to 129.70): BE0

Ion 127.00 (126.70 to 127.70): BE0

Time-->




Abundance

Ion 128.00 (127.70 to 128.70): BE0

Ion 129.00 (128.70 to 129.70): BE0

Ion 127.00 (126.70 to 127.70): BE0

Time-->



#9

2-Methylnaphthalene

Concen: 0.05 ng

RT: 12.27 min Scan# 1648

Delta R.T. -0.00 min

Lab File: BE097483.D

Acq: 18 Mar 2016 1:57

Instrument :

BNA_E

ClientSampled :

MDL-03-W


Tgt Ion:142 Resp: 1635

Ion Ratio Lower Upper

142 100

141 65.6 66.9 100.3#




115 37.1 25.1 37.7



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#10

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.44 min Scan# 2017

Delta R.T. -0.01 min

Lab File: BE097483.D

Acq: 18 Mar 2016 1:57


Tgt Ion:164 Resp: 119137

Ion Ratio Lower Upper

164 100

162 100.3 84.4 126.6



160 44.6 37.7 56.5




Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E









#11
 2,4,6-Tribromophenol
 Concen: 5.53 ng
 RT: 15.92 min Scan# 2270
 Delta R.T. -0.01 min
 Lab File: BE097483.D
 Acq: 18 Mar 2016 1:57




Instrument :
 BNA_E
 ClientSampleId :
 MDL-03-W


Tgt Ion:330 Resp: 23951
 Ion Ratio Lower Upper
 330 100
 332 91.0 79.2 118.8
 141 55.3 28.1 42.1#



#12
 2-Fluorobiphenyl
 Concen: 4.09 ng
 RT: 13.07 min Scan# 1784
 Delta R.T. -0.00 min
 Lab File: BE097483.D
 Acq: 18 Mar 2016 1:57

Tgt Ion:172 Resp: 122575
 Ion Ratio Lower Upper
 172 100
 171 36.5 29.8 44.8
 170 24.2 21.4 32.2





#13

Acenaphthylene

Concen: 0.16 ng

RT: 14.17 min Scan# 1971

Delta R.T. 0.01 min

Lab File: BE097483.D

Acq: 18 Mar 2016 1:57

Instrument :

BNA_E

ClientSampleId :

MDL-03-W


Tgt Ion:152 Resp: 1709

Ion Ratio Lower Upper

152 100

151 31.8 16.3 24.5#




153 16.3 11.1 16.7



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#14

Acenaphthene

Concen: 0.05 ng

RT: 14.50 min Scan# 2028

Delta R.T. 0.00 min

Lab File: BE097483.D

Acq: 18 Mar 2016 1:57


Tgt Ion:154 Resp: 1773

Ion Ratio Lower Upper

154 100

153 137.6 87.3 130.9#



152 75.7 42.9 64.3#




Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E





#15

Fluorene

Concen: 0.04 ng

RT: 15.49 min Scan# 2196

Delta R.T. -0.00 min

Lab File: BE097483.D

Acq: 18 Mar 2016 1:57

Instrument :

BNA_E

ClientSampleId :

MDL-03-W


Tgt Ion:166 Resp: 1985

Ion Ratio Lower Upper

166 100

165 125.2 79.2 118.8#




167 15.7 10.8 16.2



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#16

Phenanthrene-d10

Concen: 5.00 ng

RT: 17.18 min Scan# 2483

Delta R.T. -0.01 min

Lab File: BE097483.D

Acq: 18 Mar 2016 1:57


Tgt Ion:188 Resp: 292022

Ion Ratio Lower Upper

188 100

94 11.6 5.4 8.2#



80 12.2 5.0 7.4#




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BE0

Ion 80.00 (79.70 to 80.70): BE0





#17

Hexachlorobenzene

Concen: 0.06 ng

RT: 16.48 min Scan# 2365

Delta R.T. 0.00 min

Lab File: BE097483.D

Acq: 18 Mar 2016 1:57

Instrument :

BNA_E

ClientSampleId :

MDL-03-W


Tgt Ion:284 Resp: 730

Ion Ratio Lower Upper

284 100

142 62.6 31.0 46.4#




249 52.1 25.8 38.8#



Abundance Ion 283.80 (283.50 to 284.50): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E



#18

Pentachlorophenol

Concen: 0.35 ng

RT: 16.83 min Scan# 2424

Delta R.T. 0.00 min

Lab File: BE097483.D

Acq: 18 Mar 2016 1:57


Tgt Ion:266 Resp: 346

Ion Ratio Lower Upper

266 100

268 0.0 48.2 72.2#



264 18.9 52.1 78.1#




Abundance Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E





#19

Phenanthrene

Concen: 0.06 ng

RT: 17.22 min Scan# 2490

Delta R.T. -0.00 min

Lab File: BE097483.D

Acq: 18 Mar 2016 1:57

Instrument :

BNA_E

ClientSampleId :

MDL-03-W





Tgt Ion:178 Resp: 5146

Ion Ratio Lower Upper

178 100

176 29.5 15.5 23.3#

179 20.2 13.1 19.7#



#20

Anthracene

Concen: 0.05 ng

RT: 17.32 min Scan# 2507

Delta R.T. 0.01 min

Lab File: BE097483.D

Acq: 18 Mar 2016 1:57




Tgt Ion:178 Resp: 3522


Ion Ratio Lower Upper

178 100

176 29.3 14.9 22.3#

179 32.7 12.2 18.4#





#21

Fluoranthene

Concen: 0.05 ng

RT: 19.23 min Scan# 2832

Delta R.T. 0.00 min

Lab File: BE097483.D

Acq: 18 Mar 2016 1:57

Instrument :

BNA_E

ClientSampleId :

MDL-03-W


Tgt Ion:202 Resp: 4896

Ion Ratio Lower Upper

202 100

101 17.7 8.3 12.5#




203 22.8 13.9 20.9#



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#22

Chrysene-d12

Concen: 5.00 ng

RT: 21.34 min Scan# 3192

Delta R.T. 0.01 min

Lab File: BE097483.D

Acq: 18 Mar 2016 1:57


Tgt Ion:240 Resp: 347908

Ion Ratio Lower Upper

240 100

120 13.3 4.0 6.0#



236 26.4 23.0 34.4




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#23

Pyrene

Concen: 0.06 ng

RT: 19.59 min Scan# 2894

Delta R.T. 0.00 min

Lab File: BE097483.D

Acq: 18 Mar 2016 1:57

Instrument :

BNA_E

ClientSampleId :

MDL-03-W


Tgt Ion: 202 Resp: 5354

Ion Ratio Lower Upper

202 100

200 24.4 16.4 24.6




203 17.7 14.6 21.8



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#24

Terphenyl-d14

Concen: 5.32 ng

RT: 19.79 min Scan# 2928

Delta R.T. -0.01 min

Lab File: BE097483.D

Acq: 18 Mar 2016 1:57


Tgt Ion: 244 Resp: 222516

Ion Ratio Lower Upper

244 100

212 7.3 5.8 8.8



122 12.8 1.8 2.8#




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#25

Benzo(a)anthracene

Concen: 0.07 ng

RT: 21.33 min Scan# 3190

Delta R.T. 0.02 min

Lab File: BE097483.D

Acq: 18 Mar 2016 1:57

Instrument :

BNA_E

ClientSampleId :

MDL-03-W


Tgt Ion:228 Resp: 6479

Ion Ratio Lower Upper

228 100

226 27.3 24.0 36.0




229 17.6 13.8 20.6



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#26

Chrysene

Concen: Below Cal

RT: 21.38 min Scan# 3198

Delta R.T. -0.00 min

Lab File: BE097483.D

Acq: 18 Mar 2016 1:57


Tgt Ion:228 Resp: 6889

Ion Ratio Lower Upper

228 100

226 35.9 18.9 28.3#



229 21.3 19.1 28.7




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#27

Indeno(1,2,3-cd)pyrene

Concen: 0.02 ng

RT: 26.38 min Scan# 4049

Delta R.T. -0.02 min





Lab File: BE097483.D

Acq: 18 Mar 2016 1:57

Instrument :
BNA_E
ClientSampleId :
MDL-03-W

Tgt Ion:276 Resp: 2055

Ion	Ratio	Lower	Upper
276	100		
138	18.1	20.2	30.4#
277	0.0	19.8	29.8#



#28

Perylene-d12

Concen: 5.00 ng

RT: 23.80 min Scan# 3611




Delta R.T. -0.00 min


Lab File: BE097483.D

Acq: 18 Mar 2016 1:57

Tgt Ion:264 Resp: 327753

Ion	Ratio	Lower	Upper
264	100		
260	22.4	20.1	30.1
265	21.1	18.3	27.5





#29

Benzo(b)fluoranthene

Concen: 0.06 ng

RT: 23.05 min Scan# 3483

Delta R.T. 0.01 min

Lab File: BE097483.D

Acq: 18 Mar 2016 1:57





Instrument :

BNA_E

ClientSampleId :

MDL-03-W

Tgt Ion:	252	Resp:	6694
Ion Ratio	Lower	Upper	
252	100		
253	27.8	18.7	28.1
125	22.5	9.8	14.6#



#30

Benzo(k)fluoranthene

Concen: 0.05 ng




RT: 23.09 min Scan# 3490


Delta R.T. -0.01 min

Lab File: BE097483.D

Acq: 18 Mar 2016 1:57

Tgt Ion:	252	Resp:	5791
Ion Ratio	Lower	Upper	
252	100		
253	26.6	20.2	30.4
125	18.1	9.4	14.2#





#31

Benzo(a)pyrene

Concen: 0.05 ng

RT: 23.69 min Scan# 3592

Delta R.T. 0.00 min

Lab File: BE097483.D

Acq: 18 Mar 2016 1:57

Instrument :





BNA_E

ClientSampleId :

MDL-03-W

Tgt Ion:252 Resp: 4941

Ion	Ratio	Lower	Upper
252	100		
253	19.8	19.4	29.0
125	42.7	11.4	17.2#



#32

Dibenzo(a,h)anthracene

Concen: 0.04 ng

RT: 26.42 min Scan# 4056




Delta R.T. -0.00 min


Lab File: BE097483.D

Acq: 18 Mar 2016 1:57

Tgt Ion:278 Resp: 3387

Ion	Ratio	Lower	Upper
278	100		
139	50.7	14.2	21.4#
279	29.4	19.6	29.4#





#33

Benzo(g,h,i)perylene

Concen: 0.05 ng

RT: 27.19 min Scan# 4187

Delta R.T. 0.00 min

Lab File: BE097483.D

Acq: 18 Mar 2016 1:57

Instrument :

BNA_E

ClientSampleId :

MDL-03-W


Tgt Ion: 276 Resp: 4221

Ion Ratio Lower Upper

276 100

277 39.0 19.4 29.2#

138 28.7 18.2 27.4#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

