

Data Path : Z:\HPCHEM1\BNA_E\Data\BE031816\
 Data File : BE091477.D
 Acq On : 17 Mar 2016 21:37
 Operator : UM/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE031816

Quant Time: Mar 18 04:47:47 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE031816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 18 04:36:48 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	77	0.00
2	1,4-Dioxane	0.624	0.796	-27.6#	103	0.00
3	n-Nitrosodimethylamine	0.887	1.053	-18.7	105	0.00
4 S	2-Fluorophenol	1.149	1.386	-20.6	103	0.00
5 S	Phenol-d6	1.500	1.726	-15.1	102	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	76	0.00
7 S	Nitrobenzene-d5	0.224	0.268	-19.6	109	0.00
8	Naphthalene	1.221	1.509	-23.6	101	0.00
9	2-Methylnaphthalene	0.785	0.932	-18.7	98	0.00
10 I	Acenaphthene-d10	1.000	1.000	0.0	75	0.00
11 S	2,4,6-Tribromophenol	0.142	0.152	-7.0	100	0.00
12 S	2-Fluorobiphenyl	1.258	1.495	-18.8	98	0.00
13	Acenaphthylene	2.134	2.358	-10.5	102	0.00
14	Acenaphthene	1.540	1.913	-24.2	98	0.00
15	Fluorene	1.949	2.388	-22.5	99	0.00
16 I	Phenanthrene-d10	1.000	1.000	0.0	77	0.00
17	Hexachlorobenzene	0.218	0.258	-18.3	101	0.00
18	Pentachlorophenol	0.086	0.091	-5.8	100	0.00
19	Phenanthrene	1.427	1.750	-22.6	101	0.00
20	Anthracene	1.252	1.453	-16.1	102	0.00
21	Fluoranthene	1.638	1.968	-20.1	103	0.00
22 I	Chrysene-d12	1.000	1.000	0.0	83	0.00
23	Pyrene	1.244	1.450	-16.6	102	0.00
24 S	Terphenyl-d14	0.601	0.712	-18.5	100	0.00
25	Benzo(a)anthracene	1.376	1.635	-18.8	106	0.00
26	Chrysene	1.406	1.655	-17.7	107	0.00
27	Indeno(1,2,3-cd)pyrene	1.434	1.664	-16.0	106	-0.01
28 I	Perylene-d12	1.000	1.000	0.0	83	0.00
29	Benzo(b)fluoranthene	1.580	1.888	-19.5	110	0.00
30	Benzo(k)fluoranthene	1.615	1.809	-12.0	101	-0.01
31 C	Benzo(a)pyrene	1.392	1.528	-9.8	104	0.00
32	Dibenzo(a,h)anthracene	1.371	1.550	-13.1	106	0.00
33	Benzo(g,h,i)perylene	1.426	1.649	-15.6	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0