

Data Path : Z:\HPCHEM1\BNA_E\Data\BE031916\
 Data File : BE091479.D
 Acq On : 18 Mar 2016 11:30
 Operator : UM/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE031816

Quant Time: Mar 18 14:15:43 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE031816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 18 14:10:06 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	99	0.00
2	1,4-Dioxane	0.400	0.385	3.8	100	0.00
3	n-Nitrosodimethylamine	0.400	0.327	18.3	93	0.00
4 S	2-Fluorophenol	0.400	0.335	16.3	92	0.00
5 S	Phenol-d6	0.400	0.323	19.3	92	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	95	0.00
7 S	Nitrobenzene-d5	0.400	0.477	-19.2	93	0.00
8	Naphthalene	0.400	0.386	3.5	98	0.00
9	2-Methylnaphthalene	0.400	0.346	13.5	90	-0.01
10 I	Acenaphthene-d10	5.000	5.000	0.0	89	0.00
11 S	2,4,6-Tribromophenol	0.400	0.463	-15.8	79	-0.01
12 S	2-Fluorobiphenyl	0.400	0.362	9.5	88	0.00
13	Acenaphthylene	0.400	0.388	3.0	84	-0.01
14	Acenaphthene	0.400	0.362	9.5	84	0.00
15	Fluorene	0.400	0.361	9.8	86	-0.01
16 I	Phenanthrene-d10	5.000	5.000	0.0	88	0.00
17	Hexachlorobenzene	0.400	0.362	9.5	88	0.00
18	Pentachlorophenol	0.400	0.432	-8.0	85	0.00
19	Phenanthrene	0.400	0.379	5.3	89	0.00
20	Anthracene	0.400	0.341	14.8	86	0.00
21	Fluoranthene	0.400	0.347	13.3	85	0.00
22 I	Chrysene-d12	5.000	5.000	0.0	93	0.00
23	Pyrene	0.400	0.335	16.3	82	0.00
24 S	Terphenyl-d14	0.400	0.343	14.2	81	0.00
25	Benzo(a)anthracene	0.400	0.364	9.0	91	0.00
26	Chrysene	0.400	0.414	-3.5	97	-0.02
27	Indeno(1,2,3-cd)pyrene	0.400	0.341	14.8	87	-0.01
28 I	Perylene-d12	5.000	5.000	0.0	92	-0.01
29	Benzo(b)fluoranthene	0.400	0.378	5.5	96	-0.01
30	Benzo(k)fluoranthene	0.400	0.318	20.5	79	-0.01
31 C	Benzo(a)pyrene	0.400	0.323	19.3	84	-0.01
32	Dibenzo(a,h)anthracene	0.400	0.339	15.3	88	-0.01
33	Benzo(g,h,i)perylene	0.400	0.348	13.0	88	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0