

Data Path : Z:\HPCHEM1\BNA_E\Data\BE031916\
 Data File : BE091479.D
 Acq On : 18 Mar 2016 11:30
 Operator : UM/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE031816

Quant Time: Mar 18 14:15:43 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE031816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 18 14:10:06 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.624	0.600	3.8	100	0.00
3	n-Nitrosodimethylamine	0.887	0.726	18.2	93	0.00
4 S	2-Fluorophenol	1.149	0.962	16.3	92	0.00
5 S	Phenol-d6	1.500	1.212	19.2	92	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
7 S	Nitrobenzene-d5	0.224	0.184	17.9	93	0.00
8	Naphthalene	1.221	1.178	3.5	98	0.00
9	2-Methylnaphthalene	0.785	0.680	13.4	90	-0.01
10 I	Acenaphthene-d10	1.000	1.000	0.0	89	0.00
11 S	2,4,6-Tribromophenol	0.142	0.102	28.2#	79	-0.01
12 S	2-Fluorobiphenyl	1.258	1.139	9.5	88	0.00
13	Acenaphthylene	2.134	1.651	22.6	84	-0.01
14	Acenaphthene	1.540	1.396	9.4	84	0.00
15	Fluorene	1.949	1.759	9.7	86	-0.01
16 I	Phenanthrene-d10	1.000	1.000	0.0	88	0.00
17	Hexachlorobenzene	0.218	0.197	9.6	88	0.00
18	Pentachlorophenol	0.086	0.068	20.9	85	0.00
19	Phenanthrene	1.427	1.353	5.2	89	0.00
20	Anthracene	1.252	1.068	14.7	86	0.00
21	Fluoranthene	1.638	1.421	13.2	85	0.00
22 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
23	Pyrene	1.244	1.043	16.2	82	0.00
24 S	Terphenyl-d14	0.601	0.515	14.3	81	0.00
25	Benzo(a)anthracene	1.376	1.253	8.9	91	0.00
26	Chrysene	1.406	1.340	4.7	97	-0.02
27	Indeno(1,2,3-cd)pyrene	1.434	1.221	14.9	87	-0.01
28 I	Perylene-d12	1.000	1.000	0.0	92	-0.01
29	Benzo(b)fluoranthene	1.580	1.492	5.6	96	-0.01
30	Benzo(k)fluoranthene	1.615	1.284	20.5	79	-0.01
31 C	Benzo(a)pyrene	1.392	1.124	19.3	84	-0.01
32	Dibenzo(a,h)anthracene	1.371	1.162	15.2	88	-0.01
33	Benzo(g,h,i)perylene	1.426	1.242	12.9	88	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0