

Data Path : Z:\HPCHEM1\BNA_E\Data\BE031916\
 Data File : BE091480.D
 Acq On : 18 Mar 2016 12:09
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 18 14:19:40 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE031816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 18 14:14:30 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	105	0.00
2	1,4-Dioxane	0.400	0.410	-2.5	113	0.00
3	n-Nitrosodimethylamine	0.400	0.322	19.5	97	0.00
4 S	2-Fluorophenol	0.400	0.339	15.3	99	0.00
5 S	Phenol-d6	0.400	0.325	18.8	98	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	100	0.00
7 S	Nitrobenzene-d5	0.400	0.475	-18.7	97	0.00
8	Naphthalene	0.400	0.382	4.5	103	0.00
9	2-Methylnaphthalene	0.400	0.347	13.3	95	-0.01
10 I	Acenaphthene-d10	5.000	5.000	0.0	93	0.00
11 S	2,4,6-Tribromophenol	0.400	0.455	-13.7	79	-0.01
12 S	2-Fluorobiphenyl	0.400	0.367	8.3	93	0.00
13	Acenaphthylene	0.400	0.391	2.3	89	-0.01
14	Acenaphthene	0.400	0.365	8.8	88	0.00
15	Fluorene	0.400	0.366	8.5	91	-0.01
16 I	Phenanthrene-d10	5.000	5.000	0.0	92	-0.01
17	Hexachlorobenzene	0.400	0.371	7.3	94	0.00
18	Pentachlorophenol	0.400	0.408	-2.0	80	0.00
19	Phenanthrene	0.400	0.378	5.5	93	0.00
20	Anthracene	0.400	0.339	15.3	89	0.00
21	Fluoranthene	0.400	0.342	14.5	88	0.00
22 I	Chrysene-d12	5.000	5.000	0.0	95	0.00
23	Pyrene	0.400	0.337	15.8	84	0.00
24 S	Terphenyl-d14	0.400	0.330	17.5	80	0.00
25	Benzo(a)anthracene	0.400	0.355	11.3	90	0.00
26	Chrysene	0.400	0.431	-7.7	103	-0.02
27	Indeno(1,2,3-cd)pyrene	0.400	0.331	17.3	86	-0.01
28 I	Perylene-d12	5.000	5.000	0.0	91	-0.01
29	Benzo(b)fluoranthene	0.400	0.373	6.8	95	-0.01
30	Benzo(k)fluoranthene	0.400	0.323	19.3	80	-0.01
31 C	Benzo(a)pyrene	0.400	0.322	19.5	84	-0.01
32	Dibenzo(a,h)anthracene	0.400	0.339	15.3	88	-0.01
33	Benzo(g,h,i)perylene	0.400	0.345	13.8	87	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0