

Data Path : Z:\HPCHEM1\BNA_E\Data\BE031916\
 Data File : BE091480.D
 Acq On : 18 Mar 2016 12:09
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 18 14:19:40 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE031816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 18 14:14:30 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2	1,4-Dioxane	0.624	0.640	-2.6	113	0.00
3	n-Nitrosodimethylamine	0.887	0.715	19.4	97	0.00
4 S	2-Fluorophenol	1.149	0.973	15.3	99	0.00
5 S	Phenol-d6	1.500	1.218	18.8	98	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
7 S	Nitrobenzene-d5	0.224	0.182	18.8	97	0.00
8	Naphthalene	1.221	1.167	4.4	103	0.00
9	2-Methylnaphthalene	0.785	0.681	13.2	95	-0.01
10 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.00
11 S	2,4,6-Tribromophenol	0.142	0.098	31.0#	79	-0.01
12 S	2-Fluorobiphenyl	1.258	1.155	8.2	93	0.00
13	Acenaphthylene	2.134	1.666	21.9	89	-0.01
14	Acenaphthene	1.540	1.407	8.6	88	0.00
15	Fluorene	1.949	1.783	8.5	91	-0.01
16 I	Phenanthrene-d10	1.000	1.000	0.0	92	-0.01
17	Hexachlorobenzene	0.218	0.202	7.3	94	0.00
18	Pentachlorophenol	0.086	0.061	29.1#	80	0.00
19	Phenanthrene	1.427	1.348	5.5	93	0.00
20	Anthracene	1.252	1.063	15.1	89	0.00
21	Fluoranthene	1.638	1.402	14.4	88	0.00
22 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
23	Pyrene	1.244	1.047	15.8	84	0.00
24 S	Terphenyl-d14	0.601	0.496	17.5	80	0.00
25	Benzo(a)anthracene	1.376	1.222	11.2	90	0.00
26	Chrysene	1.406	1.392	1.0	103	-0.02
27	Indeno(1,2,3-cd)pyrene	1.434	1.185	17.4	86	-0.01
28 I	Perylene-d12	1.000	1.000	0.0	91	-0.01
29	Benzo(b)fluoranthene	1.580	1.475	6.6	95	-0.01
30	Benzo(k)fluoranthene	1.615	1.304	19.3	80	-0.01
31 C	Benzo(a)pyrene	1.392	1.122	19.4	84	-0.01
32	Dibenzo(a,h)anthracene	1.371	1.164	15.1	88	-0.01
33	Benzo(g,h,i)perylene	1.426	1.231	13.7	87	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0