

Data Path : Z:\HPCHEM1\BNA_E\Data\BE031916\
 Data File : BE091493.D
 Acq On : 18 Mar 2016 21:33
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 19 00:40:42 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE031816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 18 14:14:30 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	103	0.00
2	1,4-Dioxane	0.400	0.376	6.0	102	0.00
3	n-Nitrosodimethylamine	0.400	0.326	18.5	97	0.00
4 S	2-Fluorophenol	0.400	0.335	16.3	96	0.00
5 S	Phenol-d6	0.400	0.331	17.3	98	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	99	0.00
7 S	Nitrobenzene-d5	0.400	0.496	-24.0	104	0.00
8	Naphthalene	0.400	0.388	3.0	103	0.00
9	2-Methylnaphthalene	0.400	0.353	11.8	95	-0.01
10 I	Acenaphthene-d10	5.000	5.000	0.0	94	0.00
11 S	2,4,6-Tribromophenol	0.400	0.467	-16.8	84	-0.01
12 S	2-Fluorobiphenyl	0.400	0.362	9.5	92	0.00
13	Acenaphthylene	0.400	0.398	0.5	92	-0.01
14	Acenaphthene	0.400	0.367	8.3	89	0.00
15	Fluorene	0.400	0.361	9.8	90	-0.01
16 I	Phenanthrene-d10	5.000	5.000	0.0	91	-0.01
17	Hexachlorobenzene	0.400	0.368	8.0	93	0.00
18	Pentachlorophenol	0.400	0.400	0.0	76	0.00
19	Phenanthrene	0.400	0.379	5.3	93	0.00
20	Anthracene	0.400	0.341	14.8	89	-0.01
21	Fluoranthene	0.400	0.325	18.8	83	0.00
22 I	Chrysene-d12	5.000	5.000	0.0	83	0.00
23	Pyrene	0.400	0.368	8.0	80	-0.02
24 S	Terphenyl-d14	0.400	0.402	-0.5	85	-0.02
25	Benzo(a)anthracene	0.400	0.400	0.0	89	0.00
26	Chrysene	0.400	0.505	-26.2#	105	-0.02
27	Indeno(1,2,3-cd)pyrene	0.400	0.370	7.5	84	-0.02
28 I	Perylene-d12	5.000	5.000	0.0	90	-0.01
29	Benzo(b)fluoranthene	0.400	0.345	13.8	86	-0.01
30	Benzo(k)fluoranthene	0.400	0.346	13.5	85	-0.02
31 C	Benzo(a)pyrene	0.400	0.314	21.5#	80	-0.01
32	Dibenzo(a,h)anthracene	0.400	0.337	15.8	86	-0.02
33	Benzo(g,h,i)perylene	0.400	0.339	15.3	84	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 1