

Data Path : Z:\HPCHEM1\BNA_E\Data\BE031916\
 Data File : BE097493.D
 Acq On : 18 Mar 2016 21:33
 Operator : UM/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :

Quant Time: Mar 19 00:40:42 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE031816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 18 14:14:30 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.624	0.587	5.9	102	0.00
3	n-Nitrosodimethylamine	0.887	0.722	18.6	97	0.00
4 S	2-Fluorophenol	1.149	0.961	16.4	96	0.00
5 S	Phenol-d6	1.500	1.239	17.4	98	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
7 S	Nitrobenzene-d5	0.224	0.197	12.1	104	0.00
8	Naphthalene	1.221	1.184	3.0	103	0.00
9	2-Methylnaphthalene	0.785	0.693	11.7	95	-0.01
10 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
11 S	2,4,6-Tribromophenol	0.142	0.104	26.8#	84	-0.01
12 S	2-Fluorobiphenyl	1.258	1.139	9.5	92	0.00
13	Acenaphthylene	2.134	1.717	19.5	92	-0.01
14	Acenaphthene	1.540	1.414	8.2	89	0.00
15	Fluorene	1.949	1.760	9.7	90	-0.01
16 I	Phenanthrene-d10	1.000	1.000	0.0	91	-0.01
17	Hexachlorobenzene	0.218	0.200	8.3	93	0.00
18	Pentachlorophenol	0.086	0.059	31.4#	76	0.00
19	Phenanthrene	1.427	1.353	5.2	93	0.00
20	Anthracene	1.252	1.068	14.7	89	-0.01
21	Fluoranthene	1.638	1.330	18.8	83	0.00
22 I	Chrysene-d12	1.000	1.000	0.0	83	0.00
23	Pyrene	1.244	1.143	8.1	80	-0.02
24 S	Terphenyl-d14	0.601	0.605	-0.7	85	-0.02
25	Benzo(a)anthracene	1.376	1.377	-0.1	89	0.00
26	Chrysene	1.406	1.628	-15.8	105	-0.02
27	Indeno(1,2,3-cd)pyrene	1.434	1.328	7.4	84	-0.02
28 I	Perylene-d12	1.000	1.000	0.0	90	-0.01
29	Benzo(b)fluoranthene	1.580	1.363	13.7	86	-0.01
30	Benzo(k)fluoranthene	1.615	1.398	13.4	85	-0.02
31 C	Benzo(a)pyrene	1.392	1.091	21.6#	80	-0.01
32	Dibenzo(a,h)anthracene	1.371	1.156	15.7	86	-0.02
33	Benzo(g,h,i)perylene	1.426	1.210	15.1	84	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 1