

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE031919\
 Data File : BE099040.D
 Acq On : 19 Mar 2019 18:03
 Operator : JU/SJ
 Sample : K1963-08
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 BP-EB02-20190314

Quant Time: Mar 20 05:27:43 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 15 05:31:33 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	734	0.40	ng	0.00
7) Naphthalene-d8	10.60	136	3316	0.40	ng	0.01
13) Acenaphthene-d10	14.46	164	2343	0.40	ng	0.00
19) Phenanthrene-d10	17.21	188	5357	0.40	ng	0.00
27) Chrysene-d12	21.40	240	6281	0.40	ng	0.00
34) Perylene-d12	23.90	264	5645	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.42	112	331	0.15	ng	0.04
5) Phenol-d6	7.04	99	260	0.08	ng	0.06
8) Nitrobenzene-d5	8.99	82	1285	0.37	ng	0.03
11) 2-Methylnaphthalene-d10	12.21	152	2457	0.40	ng	0.01
14) 2,4,6-Tribromophenol	15.97	330	444	0.38	ng	0.00
15) 2-Fluorobiphenyl	13.09	172	4424	0.46	ng	0.00
25) Fluoranthene-d10	19.24	212	33896	0.40	ng	0.00
29) Terphenyl-d14	19.83	244	9035	0.59	ng	-0.01
Target Compounds						Qvalue
32) Bis(2-ethylhexyl)phthalate	21.29	149	2072	0.050	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE031919\
Data File : BE099040.D
Acq On : 19 Mar 2019 18:03
Operator : JU/SJ
Sample : K1963-08
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
BP-EB02-20190314

Quant Time: Mar 20 05:27:43 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 15 05:31:33 2019
Response via : Initial Calibration

