

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031919\
 Data File : BE099031.D
 Acq On : 19 Mar 2019 11:09
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 20 09:50:19 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 15 05:31:33 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	123	0.00
2	1,4-Dioxane	0.400	0.361	9.8	116	0.00
3	n-Nitrosodimethylamine	0.400	0.164	59.0#	49	0.05
4 S	2-Fluorophenol	0.400	0.314	21.5	93	0.02
5 S	Phenol-d6	0.400	0.342	14.5	105	0.01
6	bis(2-Chloroethyl)ether	0.400	0.304	24.0	97	0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	133	0.00
8 S	Nitrobenzene-d5	0.400	0.470	-17.5	173	0.01
9	Naphthalene	0.400	0.400	0.0	136	0.00
10	Hexachlorobutadiene	0.400	0.385	3.8	132	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.393	1.8	133	0.00
12	2-Methylnaphthalene	0.400	0.375	6.3	128	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	140	0.00
14 S	2,4,6-Tribromophenol	0.400	0.283	29.3#	100	0.00
15 S	2-Fluorobiphenyl	0.400	0.549	-37.3#	195	-0.01
16	Acenaphthylene	0.400	0.375	6.3	138	-0.01
17	Acenaphthene	0.400	0.391	2.3	138	-0.01
18	Fluorene	0.400	0.370	7.5	133	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	120	0.00
20	4-Bromophenyl-phenylether	0.400	0.422	-5.5	129	0.00
21	Hexachlorobenzene	0.400	0.427	-6.7	135	-0.01
22	Pentachlorophenol	0.400	0.357	10.8	64	0.01
23	Phenanthrene	0.400	0.402	-0.5	124	-0.01
24	Anthracene	0.400	0.355	11.3	113	0.00
25 SURR	Fluoranthene-d10	0.400	0.359	10.3	112	0.00
26	Fluoranthene	0.400	0.376	6.0	117	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	109	-0.01
28	Pvrene	0.400	0.426	-6.5	116	0.00
29 S	Terphenyl-d14	0.400	1.175	-193.8#	321	0.00
30	Benzo(a)anthracene	0.400	0.341	14.8	95	0.00
31	Chrysene	0.400	0.427	-6.7	119	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.327	18.3	97	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.334	16.5	92	0.00
34 I	Perylene-d12	0.400	0.400	0.0	110	-0.01
35	Benzo(b)fluoranthene	0.400	0.395	1.3	113	0.00
36	Benzo(k)fluoranthene	0.400	0.437	-9.2	127	-0.01
37 C	Benzo(a)pyrene	0.400	0.409	-2.2	117	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.287	28.3#	80	0.00
39	Benzo(a,h,i)perylene	0.400	0.329	17.8	93	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031919\
Data File : BE099031.D
Acq On : 19 Mar 2019 11:09
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 20 09:50:19 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 15 05:31:33 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				