

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031919\
 Data File : BE099031.D
 Acq On : 19 Mar 2019 11:09
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Mar 20 09:50:19 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 15 05:31:33 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	123	0.00
2	1,4-Dioxane	0.789	0.716	9.3	116	0.00
3	n-Nitrosodimethylamine	1.006	0.412	59.0#	49#	0.05
4 S	2-Fluorophenol	1.196	0.937	21.7	93	0.02
5 S	Phenol-d6	1.689	1.446	14.4	105	0.01
6	bis(2-Chloroethyl)ether	1.304	0.989	24.2	97	0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	133	0.00
8 S	Nitrobenzene-d5	0.419	0.493	-17.7	173#	0.01
9	Naphthalene	4.581	4.576	0.1	136	0.00
10	Hexachlorobutadiene	0.303	0.291	4.0	132	0.00
11 SURR	2-Methylnaphthalene-d10	0.739	0.727	1.6	133	0.00
12	2-Methylnaphthalene	0.743	0.697	6.2	128	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	140	0.00
14 S	2,4,6-Tribromophenol	0.202	0.143	29.2#	100	0.00
15 S	2-Fluorobiphenyl	1.656	2.274	-37.3#	195#	-0.01
16	Acenaphthylene	2.057	1.928	6.3	138	-0.01
17	Acenaphthene	1.220	1.193	2.2	138	-0.01
18	Fluorene	1.760	1.627	7.6	133	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	120	0.00
20	4-Bromophenyl-phenylether	0.217	0.229	-5.5	129	0.00
21	Hexachlorobenzene	0.258	0.275	-6.6	135	-0.01
22	Pentachlorophenol	0.052	0.020	61.5#	64	0.01
23	Phenanthrene	1.137	1.142	-0.4	124	-0.01
24	Anthracene	0.972	0.862	11.3	113	0.00
25 SURR	Fluoranthene-d10	6.385	5.730	10.3	112	0.00
26	Fluoranthene	1.605	1.508	6.0	117	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	109	-0.01
28	Pvrene	1.224	1.303	-6.5	116	0.00
29 S	Terphenyl-d14	0.972	2.855	-193.7#	321#	0.00
30	Benzo(a)anthracene	1.248	1.065	14.7	95	0.00
31	Chrysene	1.336	1.427	-6.8	119	-0.01
32	Bis(2-ethylhexyl)phthalate	2.657	2.171	18.3	97	-0.01
33	Indeno(1,2,3-cd)pyrene	1.410	1.178	16.5	92	0.00
34 I	Perylene-d12	1.000	1.000	0.0	110	-0.01
35	Benzo(b)fluoranthene	1.315	1.299	1.2	113	0.00
36	Benzo(k)fluoranthene	1.377	1.506	-9.4	127	-0.01
37 C	Benzo(a)pyrene	1.261	1.291	-2.4	117	-0.01
38	Dibenzo(a,h)anthracene	1.193	0.856	28.2#	80	0.00
39	Benzo(a,h,i)perylene	1.248	1.027	17.7	93	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031919\
Data File : BE099031.D
Acq On : 19 Mar 2019 11:09
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 20 09:50:19 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 15 05:31:33 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			