

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031919\
 Data File : BE099046.D
 Acq On : 19 Mar 2019 21:46
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 20 05:39:27 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 15 05:31:33 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2	1,4-Dioxane	0.789	0.761	3.5	111	0.00
3	n-Nitrosodimethylamine	1.006	0.523	48.0#	56	0.04
4 S	2-Fluorophenol	1.196	0.938	21.6	84	0.01
5 S	Phenol-d6	1.689	1.636	3.1	107	0.01
6	bis(2-Chloroethyl)ether	1.304	1.037	20.5	91	0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	127	0.00
8 S	Nitrobenzene-d5	0.419	0.310	26.0#	104	0.01
9	Naphthalene	4.581	4.525	1.2	128	-0.01
10	Hexachlorobutadiene	0.303	0.281	7.3	121	-0.01
11 SURR	2-Methylnaphthalene-d10	0.739	0.714	3.4	125	0.00
12	2-Methylnaphthalene	0.743	0.674	9.3	119	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	133	0.00
14 S	2,4,6-Tribromophenol	0.202	0.137	32.2#	91	0.00
15 S	2-Fluorobiphenyl	1.656	1.546	6.6	126	0.00
16	Acenaphthylene	2.057	1.964	4.5	133	-0.02
17	Acenaphthene	1.220	1.192	2.3	130	-0.02
18	Fluorene	1.760	1.612	8.4	125	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	113	-0.01
20	4-Bromophenyl-phenylether	0.217	0.215	0.9	114	0.00
21	Hexachlorobenzene	0.258	0.254	1.6	117	-0.01
22	Pentachlorophenol	0.052	0.027	48.1#	80	0.01
23	Phenanthrene	1.137	1.062	6.6	108	-0.01
24	Anthracene	0.972	0.810	16.7	100	0.00
25 SURR	Fluoranthene-d10	6.385	5.601	12.3	103	-0.01
26	Fluoranthene	1.605	1.392	13.3	102	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	95	-0.01
28	Pvrene	1.224	1.294	-5.7	100	0.00
29 S	Terphenyl-d14	0.972	0.936	3.7	92	0.00
30	Benzo(a)anthracene	1.248	0.984	21.2	76	0.00
31	Chrysene	1.336	1.384	-3.6	101	-0.01
32	Bis(2-ethylhexyl)phthalate	2.657	2.269	14.6	88	-0.01
33	Indeno(1,2,3-cd)pyrene	1.410	1.093	22.5	75	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	97	-0.01
35	Benzo(b)fluoranthene	1.315	1.241	5.6	95	0.00
36	Benzo(k)fluoranthene	1.377	1.313	4.6	97	-0.01
37 C	Benzo(a)pyrene	1.261	1.215	3.6	97	-0.01
38	Dibenzo(a,h)anthracene	1.193	0.824	30.9#	68	0.00
39	Benzo(a,h,i)perylene	1.248	0.962	22.9	77	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031919\
Data File : BE099046.D
Acq On : 19 Mar 2019 21:46
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 20 05:39:27 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 15 05:31:33 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			