

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031919\
 Data File : BE099046.D
 Acq On : 19 Mar 2019 21:46
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 20 05:39:27 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 15 05:31:33 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	110	0.00
2	1,4-Dioxane	0.400	0.390	2.5	111	0.00
3	n-Nitrosodimethylamine	0.400	0.208	48.0#	56	0.04
4 S	2-Fluorophenol	0.400	0.314	21.5	84	0.01
5 S	Phenol-d6	0.400	0.387	3.3	107	0.01
6	bis(2-Chloroethyl)ether	0.400	0.318	20.5	91	0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	127	0.00
8 S	Nitrobenzene-d5	0.400	0.295	26.3#	104	0.01
9	Naphthalene	0.400	0.395	1.3	128	-0.01
10	Hexachlorobutadiene	0.400	0.371	7.3	121	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.386	3.5	125	0.00
12	2-Methylnaphthalene	0.400	0.363	9.3	119	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	133	0.00
14 S	2,4,6-Tribromophenol	0.400	0.272	32.0#	91	0.00
15 S	2-Fluorobiphenyl	0.400	0.374	6.5	126	0.00
16	Acenaphthylene	0.400	0.382	4.5	133	-0.02
17	Acenaphthene	0.400	0.391	2.3	130	-0.02
18	Fluorene	0.400	0.366	8.5	125	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	113	-0.01
20	4-Bromophenyl-phenylether	0.400	0.396	1.0	114	0.00
21	Hexachlorobenzene	0.400	0.394	1.5	117	-0.01
22	Pentachlorophenol	0.400	0.388	3.0	80	0.01
23	Phenanthrene	0.400	0.374	6.5	108	-0.01
24	Anthracene	0.400	0.333	16.8	100	0.00
25 SURR	Fluoranthene-d10	0.400	0.351	12.3	103	-0.01
26	Fluoranthene	0.400	0.347	13.3	102	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	95	-0.01
28	Pvrene	0.400	0.423	-5.7	100	0.00
29 S	Terphenyl-d14	0.400	0.385	3.8	92	0.00
30	Benzo(a)anthracene	0.400	0.315	21.3	76	0.00
31	Chrysene	0.400	0.414	-3.5	101	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.342	14.5	88	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.310	22.5	75	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	97	-0.01
35	Benzo(b)fluoranthene	0.400	0.377	5.8	95	0.00
36	Benzo(k)fluoranthene	0.400	0.381	4.8	97	-0.01
37 C	Benzo(a)pyrene	0.400	0.385	3.8	97	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.276	31.0#	68	0.00
39	Benzo(a,h,i)perylene	0.400	0.308	23.0	77	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031919\
Data File : BE099046.D
Acq On : 19 Mar 2019 21:46
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 20 05:39:27 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 15 05:31:33 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				